

EX-1011152

ANALYTICAL RESULTS
FOR
U.S. GEOLOGICAL SURVEY
ENSECO-RMAL NO. 011152



SEPTEMBER 25, 1990

KAFB1047



ANALYTICAL RESULTS
FOR
U.S. GEOLOGICAL SURVEY
ENSECO-RMAL NO. 011152




SEPTEMBER 25, 1990

Reviewed by:



Randall Thompson



Lindsay Breyer

4955 Yarrow Street
Arvada, Colorado 80002
303/421-6611

Facsimile: 303/431-7171

Introduction

This report presents the analytical results as well as supporting information to aid in the evaluation and interpretation of the data and is arranged in the following order:

- o Sample Description Information
- o Analytical Test Requests
- o Analytical Results
- o Quality Control Report

Sample Description Information

The Sample Description Information lists all of the samples received in this project together with the internal laboratory identification number assigned for each sample. Each project received at Enseco - RMAL is assigned a unique six digit number. Samples within the project are numbered sequentially. The laboratory identification number is a combination of the six digit project code and the sample sequence number.

Also given in the Sample Description Information is the Sample Type (matrix), Date of Sampling (if known) and Date of Receipt at the laboratory.

Analytical Test Requests

The Analytical Test Requests lists the analyses that were performed on each sample. The Custom Test column indicates where tests have been modified to conform to the specific requirements of this project.

SAMPLE DESCRIPTION INFORMATION
for
U.S. Geological Survey

Lab ID	Client ID	Matrix	Sampled Date	Time	Received Date
011152-0001-SA	KAFB060903-2	AQUEOUS	29 AUG 90	15:30	30 AUG 90

ANALYTICAL TEST REQUESTS
for
U.S. Geological Survey

Lab ID: 011152	Group Code	Analysis Description	Custom Test?
0001	A	Total Organic Carbon (TOC)	N
		Total Organic Halogen (TOX)	N
		Nitrate Plus Nitrite	N
		Chromium, Furnace AA	N
		Chromium, Furnace AA (Total)	N
		Prep - Total Metals, ICP	N
		Appendix IX Herbicides	N
		Prep - Herbicides by GC	N

Analytical Results

The analytical results for this project are presented in the following data tables. Each data table includes sample identification information, and when available and appropriate, dates sampled, received, authorized, prepared and analyzed. The authorization data is the date when the project was defined by the client such that laboratory work could begin.

Data sheets contain a listing of the parameters measured in each test, the analytical results and the Enseco reporting limit. Reporting limits are adjusted to reflect dilution of the sample, when appropriate. Solid and waste samples are reported on an "as received" basis, i.e. no correction is made for moisture content.

Enseco-RMAL is no longer routinely blank-correcting analytical data. Uncorrected analytical results are reported, along with associated blank results, for all organic and metals analyses. Analytical results and blank results are reported for conventional inorganic parameters as specified in the method. This policy is described in detail in the Enseco Incorporated Quality Assurance Program Plan for Environmental Chemical Monitoring, Revision 3.3, April, 1989.

The results from the Standard Enseco QA/QC Program, which generates data which are independent of matrix effects, is provided subsequently.

Appendix IX Herbicides

Method 8150

Client Name: U.S. Geological Survey
 Client ID: KAFB060903-2
 Lab ID: 011152-0001-SA Enseco ID: 1090609
 Matrix: AQUEOUS Sampled: 29 AUG 90 Received: 30 AUG 90
 Authorized: 30 AUG 90 Prepared: 05 SEP 90 Analyzed: 19 SEP 90

Parameter	Result	Units	Reporting Limit	
2,4-D	ND	ug/L	1.2	T
2,4,5-TP (Silvex)	ND	ug/L	0.17	
2,4,5-T	ND	ug/L	0.20	
DCAA	109	%		

Note T : Preferred values unless footnoted on secondary column test.

ND = Not detected
 NA = Not applicable

Reported By: William Sullivan

Approved By: Jeff Lowry

Metals

Total Metals

Client Name: U.S. Geological Survey
 Client ID: KAFB060903-2
 Lab ID: 011152-0001-SA Enseco ID: 1090609
 Matrix: AQUEOUS Sampled: 29 AUG 90 Received: 30 AUG 90
 Authorized: 30 AUG 90 Prepared: See Below Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chromium	0.014	mg/L	0.010	7191	11 SEP 90	18 SEP 90

ND = Not detected
 NA = Not applicable

Reported By: Mike Befort

Approved By: Dave Roberts

Metals

Dissolved Metals

Client Name: U.S. Geological Survey

Client ID: KAFB060903-2

Lab ID: 011152-0001-SA

Matrix: AQUEOUS

Authorized: 30 AUG 90

Enseco ID: 1090609

Sampled: 29 AUG 90

Prepared: See Below

Received: 30 AUG 90

Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chromium	ND	mg/L	0.0050	7191	NA	18 SEP 90

ND = Not detected

NA = Not applicable

Reported By: Mike Befort

Approved By: Dave Roberts

General Inorganics

Client Name: U.S. Geological Survey

Client ID: KAFB060903-2

Lab ID: 011152-0001-SA

Matrix: AQUEOUS

Authorized: 30 AUG 90

Enseco ID: 1090609

Sampled: 29 AUG 90

Prepared: See Below

Received: 30 AUG 90

Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Nitrate plus Nitrite	20.9	mg/L	1.0	353.2	NA	06 SEP 90
Total Organic Carbon	0.93	mg/L	0.50	9060	NA	31 AUG 90
Total Organic Halogen as Cl	ND	ug/L	30.0	9020	NA	10 SEP 90

ND = Not detected
NA = Not applicable

Reported By: Steve Pope

Approved By: Toni Stovall

Quality Control Results

The Enseco laboratories operate under a vigorous QA/QC program designed to ensure the generation of scientifically valid, legally defensible data by monitoring every aspect of laboratory operations. Routine QA/QC procedures include the use of approved methodologies, independent verification of analytical standards, use of duplicate Laboratory Control Samples to assess the precision and accuracy of the methodology on a routine basis, and a rigorous system of data review.

In addition, the Enseco laboratories maintain a comprehensive set of certifications from both state and federal governmental agencies which require frequent analyses of blind audit samples. Enseco - Rocky Mountain Analytical Laboratory is certified by the EPA under the EPA/CLP program for both Organic and Inorganic analyses, under the USATHAMA (U.S. Army) program, by the Army Corps of Engineers, and the states of Colorado, New Jersey, New York, Utah, and Florida, among others.

The standard laboratory QC package is designed to:

- 1) establish a strong, cost-effective QC program that ensures the generation of scientifically valid, legally defensible data
- 2) assess the laboratory's performance of the analytical method using control limits generated with a well-defined matrix
- 3) establish clear-cut guidelines for acceptability of analytical data so that QC decisions can be made immediately at the bench, and
- 4) provide a standard set of reportables which assures the client of the quality of his data.

The Enseco QC program is based upon monitoring the precision and accuracy of an analytical method by analyzing a set of Duplicate Control Samples (DCS) at frequent, well-defined intervals. Each DCS is a well-characterized matrix which is spiked with target compounds at 5-100 times the reporting limit, depending upon the methodology being monitored. The purpose of the DCS is not to duplicate the sample matrix, but rather to provide an interference-free, homogeneous matrix from which to gather data to establish control limits. These limits are used to determine whether data generated by the laboratory on any given day is in control.

Control limits for accuracy (percent recovery) are based on the average, historical percent recovery +/- 3 standard deviation units. Control limits for precision (relative percent difference) range from 0 (identical duplicate DCS results) to the average, historical relative percent difference + 3 standard deviation units. These control limits are fairly narrow based on the consistency of the matrix being monitored and are updated on a quarterly basis.

For each batch of samples analyzed, an additional control measure is taken in the form of a Single Control Sample (SCS). The SCS consists of a control matrix that is spiked with either representative target compounds or surrogate compounds appropriate to the method being used. An SCS is prepared for each sample lot for which the DCS pair are not analyzed.

Accuracy for DCS and SCS is measured by Percent Recovery.

$$\% \text{ Recovery} = \frac{\text{Measured Concentration}}{\text{Actual Concentration}} \times 100$$

Precision for DCS is measured by Relative Percent Difference (RPD).

$$\text{RPD} = \frac{|\text{Measured Concentration DCS1} - \text{Measured Concentration DCS2}|}{(\text{Measured Concentration DCS1} + \text{Measured Concentration DCS2})/2} \times 100$$

All samples analyzed concurrently by the same test are assigned the same QC lot number. Projects which contain numerous samples, analyzed over several days, may have multiple QC lot numbers associated with each test. The QC information which follows includes a listing of the QC lot numbers associated with each of the samples reported, DCS and SCS (where applicable) recoveries from the QC lots associated with the samples, and control limits for these lots. The QC data is reported by test code, in the order that the tests are reported in the analytical results section of this report.

QC LOT ASSIGNMENT REPORT
Semivolatile Organics by GC

Laboratory Sample Number	QC Matrix	QC Category	QC Lot Number (DCS)	QC Run Number (SCS/BLANK)
011152-0001-SA	AQUEOUS	615-A	05 SEP 90-A	05 SEP 90-A

DUPLICATE CONTROL SAMPLE REPORT
Semivolatile Organics by GC

Analyte	Concentration Spiked	Concentration Measured		AVG	Accuracy Average (%)		Precision (RPD)		
		DCS1	DCS2		DCS	Limits	DCS	Limit	
Category: 615-A									
Matrix: AQUEOUS									
QC Lot: 05 SEP 90-A									
Concentration Units: ug/L									
2,4-D	5.0	4.75	4.71	4.73	95	19-129	0.9	54	
2,4,5-TP (Silvex)	1.0	0.922	0.919	0.920	92	23-127	0.3	39	
2,4,5-T	1.0	1.04	1.03	1.04	104	40-112	1.0	20	

Calculations are performed before rounding to avoid round-off errors in calculated results.

SINGLE CONTROL SAMPLE REPORT
Semivolatile Organics by GC

Analyte	Concentration		Accuracy(%)	
	Spiked	Measured	SCS	Limits
Category: 615-A				
Matrix: AQUEOUS				
QC Lot: 05 SEP 90-A QC Run: 05 SEP 90-A				
Concentration Units: ug/L				
DCAA	5.00	6.02	120	60-120

Calculations are performed before rounding to avoid round-off errors in calculated results.

METHOD BLANK REPORT
Semivolatile Organics by GC

Analyte	Result	Units	Reporting Limit
Test: 8150-AP9-A			
Matrix: AQUEOUS			
QC Lot: 05 SEP 90-A QC Run: 05 SEP 90-A			
2,4-D	ND	ug/L	1.2
2,4,5-TP (Silvex)	ND	ug/L	0.17
2,4,5-T	ND	ug/L	0.20

QC LOT ASSIGNMENT REPORT
Metals Analysis and Preparation

Laboratory Sample Number	QC Matrix	QC Category	QC Lot Number (DCS)	QC Run Number (SCS/BLANK)
011152-0001-SA	AQUEOUS	CR-FAA-AD	18 SEP 90-A	-
011152-0001-SA	AQUEOUS	CR-FAA-AT	11 SEP 90-A	11 SEP 90-A

DUPLICATE CONTROL SAMPLE REPORT
Metals Analysis and Preparation

Analyte	Spiked	Concentration		AVG	Accuracy		Precision	
		DCS1	Measured DCS2		DCS	Average (%) Limits	(RPD) DCS	Limit

Category: CR-FAA-AD
Matrix: AQUEOUS
QC Lot: 18 SEP 90-A
Concentration Units: mg/L

Chromium	0.02	0.0188	0.0185	0.0186	93	75-125	1.6	20
----------	------	--------	--------	--------	----	--------	-----	----

Category: CR-FAA-AT
Matrix: AQUEOUS
QC Lot: 11 SEP 90-A
Concentration Units: mg/L

Chromium	0.20	0.213	0.217	0.215	107	75-125	1.9	20
----------	------	-------	-------	-------	-----	--------	-----	----

Calculations are performed before rounding to avoid round-off errors in calculated results.

METHOD BLANK REPORT
Metals Analysis and Preparation

Analyte	Result	Units	Reporting Limit
Test: CR-FAA-AT			
Matrix: AQUEOUS			
QC Lot: 11 SEP 90-A QC Run: 11 SEP 90-A			
Chromium	ND	mg/L	0.0050

QC LOT ASSIGNMENT REPORT
Wet Chemistry Analysis and Preparation

Laboratory Sample Number	QC Matrix	QC Category	QC Lot Number (DCS)	QC Run Number (SCS/BLANK)
011152-0001-SA	AQUEOUS	TOC-A	31 AUG 90-A	-
011152-0001-SA	AQUEOUS	TOX-A	10 SEP 90-B	-
011152-0001-SA	AQUEOUS	NO3-A	06 SEP 90-B	-

DUPLICATE CONTROL SAMPLE REPORT
Wet Chemistry Analysis and Preparation

Analyte	Spiked	Concentration		AVG	Accuracy		Precision		
		DCS1	Measured DCS2		Average (%) DCS	Limits	(RPD) DCS	Limit	
Category: TOC-A Matrix: AQUEOUS QC Lot: 31 AUG 90-A Concentration Units: mg/L									
Total Organic Carbon	25	25.8	25.9	25.8	103	91-109	0.4	20	
Category: TOX-A Matrix: AQUEOUS QC Lot: 10 SEP 90-B Concentration Units: ug Cl/L									
Total Organic Halogen as Cl	100	105	99.2	102	102	80-120	5.7	20	
Category: NO3-A Matrix: AQUEOUS QC Lot: 06 SEP 90-B Concentration Units: mg/L									
Nitrate as N	5.4	5.40	5.59	5.50	102	91-109	3.5	10	

Calculations are performed before rounding to avoid round-off errors in calculated results.



Appendix

Special Handling

(Circle as appropriate and explain in record 5)

Hazardous material

Site Type (circle one)

SW - Surface Water
GW - Ground Water
ME - Meteorological

LK - Lake
ES - Estuary
SP - Spring
SS - Special Source

KAFB060903-2

Station Name

USGS/WRD/NEW MEXICO SWMU

Field Office

Project

USGS

Collector

(505) 262-6678

Phone (FTS)

BILL DAM

File Deposition*

(Circle one)

Q - WATSTORE

X - Lab File

Sample identification

KAFB060903-2

Station ID or Unique Number*

463536001

Project Account #

1990
Year*

08
Month*

29
Day*

1530
Time*

08
Month

29
Day

1630
Time

NM
State Code*

035
District/ User Code*

001
County Code

Begin Date

Composite End Date

6=SW
9=SW

6
Sample Medium**

Geologic Unit

H or 9
Analysis Status**

G
Analysis Source**

9
Hydrologic Condition**

9
Sample Type**

9
Hydrologic Event**

Analysis level codes and schedules

I = SPIKE
S = DUPLICATE
9 = REGULAR

J = STORM
9 = ROUTINE

PARAMETER:

METHOD:

TOC
SW 9060

TOX
SW 9020

NO2+NO3
E 353.2

DISSOLVED CHROMIUM
SW 3005/SW 791

TOTAL CHROMIUM
SW 3020/SW 791

PARAMETER:

METHOD:

~~TOX~~
~~SW 5030/SW 6010~~

~~VOC~~
~~SW 5030/SW 6010~~

~~BIOXIN~~
~~SW 5520/SW 8280~~

~~HERBICIDES~~
~~SW 3520/SW 8150~~

PARAMETER:

METHOD:

~~APPENDIX IX ICP DISSOLVED
Co, Cu, Fe, Pb, Mn, Mo, Ni, V, Zn~~
~~SW 3005/SW 6010~~

~~APPENDIX IX ICP TOTAL
Co, Cu, Fe, Pb, Mn, Mo, Ni, V, Zn~~
~~SW 3010/SW 6010~~

PARAMETER:

METHOD:

~~APP IX SEM-VOC~~
~~SW 8280/SW 8270~~

Chain-of-Custody Record

PROJECT NAME KIRTLAND AFB SWMU PROJECT NO. 463536001 P.O. NO.

Relinquished by: (Signature) Received by: (Signature) Date Time

Relinquished by: (Signature) Received by: AIRBORNE EXPRESS Date Time 1815

Relinquished by: (Signature) Received by: (Signature) Date Time

Relinquished by: (Signature) Received at lab by: (Signature) Date Time

Relinquished from lab by: (Signature) Received by: (Signature) Date Time 08-30-90 0900

Comments (Only 50 characters stored in NWIS)

Record 5

Record 6

Total number of sample bottles for this request: 5 + 1 EXTRA GCC SHIP TO:

Enseco-Rocky Mountain Analytical
4955 Yarrow Street
Arvada, CO 80002
(303) 421-6611

Special Handling (Circle as appropriate and explain in record 5)

Site Type (circle one)

Hazardous material

SW - Surface Water
GW - Ground Water
ME - Meteorological

LK - Lake
ES - Estuary
SP - Spring
SS - Special Source

KAFB060903-2 Station Name

USGS/WRD/NEW MEXICO SWMU Field Office Project

USGS Collector (505) 262-6678 Phone (FTS) BILL DAM

File Deposition* (Circle one)

Sample identification

Q - WATSTORE
X - Lab File

For Laboratory Use Only

KAFB060903-2 Station ID or Unique Number*

463536001 Project Account #

1990 08 29 1530 08 29 1630 NM 035 001
Year* Month* Day* Time* Month Day Time State Code* District/ User Code* County Code

6=GW
9=SW

Analysis level codes and schedules

1=SPIKE
5=DUPLICATE
9=REGULAR
J=STORM
9=ROUTINE

6 Sample Medium**
Geologic Unit
H or 9 Analysis Status**
G Analysis Source**
9 Hydrologic Condition**
9 Sample Type**
9 Hydrologic Event**

PARAMETER: TOC TOX NO2+NO3 DISSOLVED CHROMIUM TOTAL CHROMIUM
METHOD: SW9060 SW9020 E353.2 SW3005/SW7191 SW3020/SW791
PARAMETER: ~~VOC~~ ~~APP. IX~~ ~~BIOXIN~~ ~~APP. IX~~
METHOD: ~~SW5030/SW6010~~ ~~SW5030/SW8240~~ ~~SW3520/SW8280~~ ~~SW3520/SW8150~~
PARAMETER: ~~APP. IX ICP DISSOLVED~~ ~~APP. IX ICP TOTAL~~
METHOD: ~~SW3005/SW6010~~ ~~SW3010/SW6010~~
PARAMETER: ~~APP. IX SEMI-VOC~~
METHOD: ~~SW8220/SW8270~~

Chain-of-Custody Record

PROJECT NAME KIRTLAND AFB SWMU PROJECT NO. 463536001 P.O. NO.

Relinquished by: (Signature) Received by: (Signature) Date Time
AIRBORNE EXPRESS 1815
Relinquished by: (Signature) Received by: (Signature) Date Time
Relinquished by: (Signature) Received at lab by: (Signature) Date Time
08-30-90 0800
Relinquished from lab by: (Signature) Received by: (Signature) Date Time

Comments (Only 50 characters stored in NWIS)

Record 5
Record 6

Total number of sample bottles for this request: 5 + 1 EXTRA GCC SHIP TO:

Enseco-Rocky Mountain Analytical
4955 Yarrow Street
Arvada, CO 80002
(303) 421-6611

ATTENTION: THOMPSON, BREYER, OR MCDEVITT