

5 of 10
C

ENTERED

ANALYTICAL RESULTS
FOR
U.S. GEOLOGICAL SURVEY
ENSECO-RMAL NO. 015216

Enseco

JUNE 29, 1991

Reviewed by:


Randall Thompson


Lindsay Breyer

Enseco Incorporated
4955 Yarrow Street
Arvada, Colorado 80002
303/421-6611 Fax: 303/431-7171

KAFB1141



Introduction

This report presents the analytical results as well as supporting information to aid in the evaluation and interpretation of the data and is arranged in the following order:

- o Sample Description Information
- o Analytical Test Requests
- o Analytical Results
- o Quality Control Report

As directed by Bill Dam, chloride has been added to the analysis of samples 015216-0002 and -0003.

Sample Description Information

The Sample Description Information lists all of the samples received in this project together with the internal laboratory identification number assigned for each sample. Each project received at Enseco - RMAL is assigned a unique six digit number. Samples within the project are numbered sequentially. The laboratory identification number is a combination of the six digit project code and the sample sequence number.

Also given in the Sample Description Information is the Sample Type (matrix), Date of Sampling (if known) and Date of Receipt at the laboratory.

Analytical Test Requests

The Analytical Test Requests lists the analyses that were performed on each sample. The Custom Test column indicates where tests have been modified to conform to the specific requirements of this project.

SAMPLE DESCRIPTION INFORMATION
for
U.S. Geological Survey

Lab ID	Client ID	Matrix	Sampled Date	Time	Received Date
015216-0001-SA	MVMWK07-2	AQUEOUS	31 MAY 91	12:17	01 JUN 91
015216-0002-SA	KAFB060808-2	AQUEOUS	31 MAY 91	13:50	01 JUN 91
015216-0003-SA	KAFB060809-2	AQUEOUS	31 MAY 91	14:55	01 JUN 91

ANALYTICAL TEST REQUESTS
for
U.S. Geological Survey

Lab ID: 015216	Group Code	Analysis Description	Custom Test?
0001	A	Nitrate Plus Nitrite	N
0002 - 0003	B	Chromium, Furnace AA (Total)	N
		Prep - Total Metals, ICP	N
		Chromium, Furnace AA	N
		Chromium VI (Total)	N
		Chromium VI (Dissolved)	N
		Nitrate Plus Nitrite	N
		Total Organic Carbon (TOC)	N
		Total Organic Halogen (TOX)	N
		Gross Alpha & Beta	N
		Uranium, Natural	N

Analytical Results

The analytical results for this project are presented in the following data tables. Each data table includes sample identification information, and when available and appropriate, dates sampled, received, authorized, prepared and analyzed. The authorization data is the date when the project was defined by the client such that laboratory work could begin.

Data sheets contain a listing of the parameters measured in each test, the analytical results and the Enseco reporting limit. Reporting limits are adjusted to reflect dilution of the sample, when appropriate. Solid and waste samples are reported on an "as received" basis, i.e. no correction is made for moisture content.

Enseco-RMAL is no longer routinely blank-correcting analytical data. Uncorrected analytical results are reported, along with associated blank results, for all organic and metals analyses. Analytical results and blank results are reported for conventional inorganic parameters as specified in the method. This policy is described in detail in the Enseco Incorporated Quality Assurance Program Plan for Environmental Chemical Monitoring, Revision 3.3, May, 1989.

The results from the Standard Enseco QA/QC Program, which generates data which are independent of matrix effects, is provided subsequently.

Metals

Total Metals

Client Name: U.S. Geological Survey
 Client ID: KAFB060808-2
 Lab ID: 015216-0002-SA
 Matrix: AQUEOUS
 Authorized: 01 JUN 91

Sampled: 31 MAY 91
 Prepared: See Below

Received: 01 JUN 91
 Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chromium (VI)	0.011	mg/L	0.010	7196	NA	01 JUN 91
Chromium	0.011	mg/L	0.0020	7191	05 JUN 91	12 JUN 91

ND = Not detected
 NA = Not applicable

Reported By: Jeff Malecha

Approved By: Toni Stovall

Metals

Total Metals

Client Name: U.S. Geological Survey
 Client ID: KAFB060809-2
 Lab ID: 015216-0003-SA
 Matrix: AQUEOUS
 Authorized: 01 JUN 91

Sampled: 31 MAY 91
 Prepared: See Below

Received: 01 JUN 91
 Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chromium (VI)	ND	mg/L	0.010	7196	NA	01 JUN 91
Chromium	ND	mg/L	0.0020	7191	05 JUN 91	12 JUN 91

ND = Not detected
 NA = Not applicable

Reported By: Jeff Malecha

Approved By: Toni Stovall

Metals

Dissolved Metals

Client Name: U.S. Geological Survey

Client ID: KAFB060808-2

Lab ID: 015216-0002-SA

Matrix: AQUEOUS

Authorized: 01 JUN 91

Sampled: 31 MAY 91

Prepared: See Below

Received: 01 JUN 91

Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chromium (VI)	ND	mg/L	0.010	7196	NA	01 JUN 91
Chromium	0.0084	mg/L	0.0020	7191	NA	06 JUN 91

ND = Not detected

NA = Not applicable

Reported By: Norma Baier

Approved By: Toni Stovall

Metals

Dissolved Metals

Client Name: U.S. Geological Survey
 Client ID: KAFB060809-2
 Lab ID: 015216-0003-SA
 Matrix: AQUEOUS
 Authorized: 01 JUN 91

Sampled: 31 MAY 91
 Prepared: See Below

Received: 01 JUN 91
 Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chromium (VI)	ND	mg/L	0.010	7196	NA	01 JUN 91
Chromium	ND	mg/L	0.0020	7191	NA	06 JUN 91

ND = Not detected
 NA = Not applicable

Reported By: Norma Baier

Approved By: Toni Stovall

General Inorganics

Enseco
A Corning Company

Client Name: U.S. Geological Survey

Client ID: MVMWK07-2

Lab ID: 015216-0001-SA

Matrix: AQUEOUS

Authorized: 01 JUN 91

Sampled: 31 MAY 91

Prepared: See Below

Received: 01 JUN 91

Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Nitrate plus Nitrite	32.3	mg/L	5.0	353.2	NA	04 JUN 91

ND = Not detected

NA = Not applicable

Reported By: Mike Settell

Approved By: Roxanne Sullivan

General Inorganics

Enseco
A Corning Company

Client Name: U.S. Geological Survey
Client ID: KAFB060808-2
Lab ID: 015216-0002-SA
Matrix: AQUEOUS
Authorized: 01 JUN 91

Sampled: 31 MAY 91
Prepared: See Below

Received: 01 JUN 91
Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chloride	27.0	mg/L	0.50	A429	NA	16 JUN 91
Nitrate plus Nitrite	32.3	mg/L	5.0	353.2	NA	04 JUN 91
Total Organic Carbon	0.62	mg/L	0.50	9060	NA	07 JUN 91
Total Organic Halogen as Cl	ND	ug/L	30.0	9020	NA	05 JUN 91

ND = Not detected
NA = Not applicable

Reported By: Tammy Bailey

Approved By: Roxanne Sullivan

General Inorganics

Client Name: U.S. Geological Survey

Client ID: KAFB060809-2

Lab ID: 015216-0003-SA

Matrix: AQUEOUS

Authorized: 01 JUN 91

Sampled: 31 MAY 91

Prepared: See Below

Received: 01 JUN 91

Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Chloride	0.56	mg/L	0.50	A429	NA	16 JUN 91
Nitrate plus Nitrite	ND	mg/L	0.050	353.2	NA	04 JUN 91
Total Organic Carbon	14.6	mg/L	0.50	9060	NA	07 JUN 91
Total Organic Halogen as Cl	ND	ug/L	30.0	9020	NA	05 JUN 91

ND = Not detected
NA = Not applicable

Reported By: Tammy Bailey

Approved By: Roxanne Sullivan

Radiochemistry

Enseco
A Corning Company

Client Name: U.S. Geological Survey
Client ID: KAFB060808-2
Lab ID: 015216-0002-SA
Matrix: AQUEOUS
Authorized: 01 JUN 91

Sampled: 31 MAY 91
Prepared: See Below

Received: 01 JUN 91
Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Gross Alpha	9.7	pCi/L	+/- 4.4	900.0	NA	17 JUN 91
Gross Beta	7.3	pCi/L	+/- 3.9	900.0	NA	17 JUN 91
Uranium, Natural	0.003	mg/L	--	ASTM D2907-70T	NA	07 JUN 91

ND = Not detected
NA = Not applicable

Reported By: Paula Hubble

Approved By: Toni Stovall

Radiochemistry

Client Name: U.S. Geological Survey
 Client ID: KAFB060809-2
 Lab ID: 015216-0003-SA
 Matrix: AQUEOUS
 Authorized: 01 JUN 91

Sampled: 31 MAY 91
 Prepared: See Below

Received: 01 JUN 91
 Analyzed: See Below

Parameter	Result	Units	Reporting Limit	Analytical Method	Prepared Date	Analyzed Date
Gross Alpha	1.3	pCi/L	+/- 1.8	900.0	NA	17 JUN 91
Gross Beta	33	pCi/L	+/- 6	900.0	NA	17 JUN 91
Uranium, Natural	<0.002	mg/L	--	ASTM D2907-70T	NA	07 JUN 91

ND = Not detected
 NA = Not applicable

Reported By: Paula Hubble

Approved By: Toni Stovall

Quality Control Results

The Enseco laboratories operate under a vigorous QA/QC program designed to ensure the generation of scientifically valid, legally defensible data by monitoring every aspect of laboratory operations. Routine QA/QC procedures include the use of approved methodologies, independent verification of analytical standards, use of duplicate Laboratory Control Samples to assess the precision and accuracy of the methodology on a routine basis, and a rigorous system of data review.

In addition, the Enseco laboratories maintain a comprehensive set of certifications from both state and federal governmental agencies which require frequent analyses of blind audit samples. Enseco - Rocky Mountain Analytical Laboratory is certified by the EPA under the EPA/CLP program for both Organic and Inorganic analyses, under the USATHAMA (U.S. Army) program, by the Army Corps of Engineers, and the states of Colorado, New Jersey, New York, Utah, and Florida, among others.

The standard laboratory QC package is designed to:

- 1) establish a strong, cost-effective QC program that ensures the generation of scientifically valid, legally defensible data
- 2) assess the laboratory's performance of the analytical method using control limits generated with a well-defined matrix
- 3) establish clear-cut guidelines for acceptability of analytical data so that QC decisions can be made immediately at the bench, and
- 4) provide a standard set of reportables which assures the client of the quality of his data.

The Enseco QC program is based upon monitoring the precision and accuracy of an analytical method by analyzing a set of Duplicate Control Samples (DCS) at frequent, well-defined intervals. Each DCS is a well-characterized matrix which is spiked with target compounds at 5-100 times the reporting limit, depending upon the methodology being monitored. The purpose of the DCS is not to duplicate the sample matrix, but rather to provide an interference-free, homogeneous matrix from which to gather data to establish control limits. These limits are used to determine whether data generated by the laboratory on any given day is in control.

Control limits for accuracy (percent recovery) are based on the average, historical percent recovery +/- 3 standard deviation units. Control limits for precision (relative percent difference) range from 0 (identical duplicate DCS results) to the average, historical relative percent difference + 3 standard deviation units. These control limits are fairly narrow based on the consistency of the matrix being monitored and are updated on a quarterly basis.

For each batch of samples analyzed, an additional control measure is taken in the form of a Single Control Sample (SCS). The SCS consists of a control matrix that is spiked with either representative target compounds or surrogate compounds appropriate to the method being used. An SCS is prepared for each sample lot for which the DCS pair are not analyzed.

Accuracy for DCS and SCS is measured by Percent Recovery.

$$\% \text{ Recovery} = \frac{\text{Measured Concentration}}{\text{Actual Concentration}} \times 100$$

Precision for DCS is measured by Relative Percent Difference (RPD).

$$\text{RPD} = \frac{|\text{Measured Concentration DCS1} - \text{Measured Concentration DCS2}|}{(\text{Measured Concentration DCS1} + \text{Measured Concentration DCS2})/2} \times 100$$

All samples analyzed concurrently by the same test are assigned the same QC lot number. Projects which contain numerous samples, analyzed over several days, may have multiple QC lot numbers associated with each test. The QC information which follows includes a listing of the QC lot numbers associated with each of the samples reported, DCS and SCS (where applicable) recoveries from the QC lots associated with the samples, and control limits for these lots. The QC data is reported by test code, in the order that the tests are reported in the analytical results section of this report.

QC LOT ASSIGNMENT REPORT
Metals Analysis and Preparation

Laboratory Sample Number	QC Matrix	QC Category	QC Lot Number (DCS)	QC Run Number (SCS/BLANK)
015216-0002-SA	AQUEOUS	CR-FAA-AT	05 JUN 91-C	05 JUN 91-C
015216-0002-SA	AQUEOUS	CR-FAA-AD	06 JUN 91-A	-
015216-0002-SA	AQUEOUS	CR6-AT	01 JUN 91-A	-
015216-0002-SA	AQUEOUS	CR6-A	01 JUN 91-A	-
015216-0003-SA	AQUEOUS	CR-FAA-AT	05 JUN 91-C	05 JUN 91-C
015216-0003-SA	AQUEOUS	CR-FAA-AD	06 JUN 91-A	-
015216-0003-SA	AQUEOUS	CR6-AT	01 JUN 91-A	-
015216-0003-SA	AQUEOUS	CR6-A	01 JUN 91-A	-

DUPLICATE CONTROL SAMPLE REPORT
Metals Analysis and Preparation

Analyte	Spiked	Concentration		AVG	Accuracy		Precision		
		DCS1	Measured DCS2		DCS	Average(%) Limits	(RPD)	DCS Limits	
Category: CR-FAA-AT Matrix: AQUEOUS QC Lot: 05 JUN 91-C Concentration Units: mg/L									
Chromium	0.20	0.211	0.210	0.210	105	75-125	0.5	20	
Category: CR-FAA-AD Matrix: AQUEOUS QC Lot: 06 JUN 91-A Concentration Units: mg/L									
Chromium	0.02	0.0199	0.0224	0.0212	106	75-125	12	20	
Category: CR6-AT Matrix: AQUEOUS QC Lot: 01 JUN 91-A Concentration Units: mg/L									
Chromium (VI)	0.05	0.0537	0.0486	0.0512	102	75-125	10	20	
Category: CR6-A Matrix: AQUEOUS QC Lot: 01 JUN 91-A Concentration Units: mg/L									
Chromium (VI)	0.050	0.0537	0.0486	0.0512	102	75-125	10	20	

Calculations are performed before rounding to avoid round-off errors in calculated results.

METHOD BLANK REPORT
Metals Analysis and Preparation

Analyte	Result	Units	Reporting Limit
Test: CR-FAA-AT			
Matrix: AQUEOUS			
QC Lot: 05 JUN 91-C	QC Run: 05 JUN 91-C		
Chromium	ND	mg/L	0.0050

QC LOT ASSIGNMENT REPORT
Wet Chemistry Analysis and Preparation

Laboratory Sample Number	QC Matrix	QC Category	QC Lot Number (DCS)	QC Run Number (SCS/BLANK)
015216-0001-SA	AQUEOUS	NO3-A	04 JUN 91-E	-
015216-0002-SA	AQUEOUS	NO3-A	04 JUN 91-E	-
015216-0002-SA	AQUEOUS	TOC-A	07 JUN 91-B	-
015216-0002-SA	AQUEOUS	TOX-A	05 JUN 91-A	-
015216-0002-SA	AQUEOUS	CL-IC-A	16 JUN 91-N	-
015216-0003-SA	AQUEOUS	NO3-A	04 JUN 91-E	-
015216-0003-SA	AQUEOUS	TOC-A	07 JUN 91-B	-
015216-0003-SA	AQUEOUS	TOX-A	05 JUN 91-A	-
015216-0003-SA	AQUEOUS	CL-IC-A	16 JUN 91-N	-

DUPLICATE CONTROL SAMPLE REPORT
Wet Chemistry Analysis and Preparation

Analyte	Spiked	Concentration		AVG	Accuracy		Precision		
		DCS1	Measured DCS2		DCS	Average (%) Limits	(RPD) DCS Limit	DCS Limit	
Category: NO3-A Matrix: AQUEOUS QC Lot: 04 JUN 91-E Concentration Units: mg/L									
Nitrate as N	7.1	7.50	7.48	7.49	105	91-109	0.3	10	
Category: TOC-A Matrix: AQUEOUS QC Lot: 07 JUN 91-B Concentration Units: mg/L									
Total Organic Carbon	25	24.8	24.8	24.8	99	91-109	0.0	20	
Category: TOX-A Matrix: AQUEOUS QC Lot: 05 JUN 91-A Concentration Units: ug Cl/L									
Total Organic Halogen as Cl	100	98.5	96.5	97.5	98	80-120	2.1	20	
Category: CL-IC-A Matrix: AQUEOUS QC Lot: 16 JUN 91-N Concentration Units: mg/L									
Chloride	20.0	20.0	20.4	20.2	101	92-108	2.0	20	

Calculations are performed before rounding to avoid round-off errors in calculated results.

Appendix

ENSECO ANALYTICAL SERVICES REQUEST FORM

15216 01

Special Handling (Circle as appropriate and explain in record 5)

Hazardous material
Sample

MVMWK07-2
Station Name

Field ID
USGS/WRD/NEW MEX
Field Office

KIRTLAND AFB
IRP-SWMU'S
Project

SW - Surface Water
GW - Ground Water
ME - Meteorological

Site Type (circle one)

LK - Lake
ES - Estuary
SP - Spring
SS - Special Source
(505) 262-5351
Phone (FTS)

File Deposition*
Circle one)

Q - WATSTORE
X - Lab File

Sample identification

For Laboratory Use Only

MVMWK07-2
Station ID or Unique Number*

463 53 600 1
Project Account #

9 9 1 05 31 12 17
Year* Month* Day* Time*
Begin Date

Month Day Time
Composite End Date

N M 0 3 5 0 0 1
State Code* District/ User Code* County Code

Analysis level codes and schedules

PARAMETER:	6 Sample Medium**	Geologic Unit	(H) or 9 Analysis Status**	9 Analysis Source**	Hydrologic Condition**	9 Sample Type**	9 Hydrologic Event**
PARAMETER:	CHROMIUM, TOTAL		CHROMIUM, DISS		CHROMIUM HEXAVALENT TOTAL	CHROMIUM HEXAVALENT DISS.	NITRATE & NITRITE
METHOD:	SW3020/SW7191		SW3005/SW7191		SW7196	SW7196	E353.2
PARAMETER:	URANIUM, GROSS ALPHA & BETA		VOC		VOX		TOC/TOX
METHOD:	AZ11B, E900		SW5030/SW8240		SW5030/SW8010		E415.1/SW9020
PARAMETER:							
METHOD:							

Chain-of-Custody Record

PROJECT NAME KIRTLAND AFB-IRP, SWMU'S PROJECT NO. 463536001 P.O. NO.

Relinquished by: (Signature) Mike Roybal	Received by: (Signature) FEDERAL EXPRESS	Date 5/31/91	Time 1630
Relinquished by: (Signature)	Received by: (Signature)	Date	Time
Relinquished by: (Signature)	Received at lab by: (Signature) Jill Man	Date 6/1/91	Time 0800
Relinquished from lab by: (Signature)	Received by: (Signature)	Date	Time

Comments (Only 50 characters stored in NWIS)

Record 5: EID WELL-SAMPLE
Record 6:

Total number of sample bottles for this request: 1

SHIP TO:

Enseco-Rocky Mountain Analytical
4955 Yarrow Street
Arvada, CO 80002
(303) 421-6611
LINDSAY BREYER

ATTENTION:

ENSECO ANALYTICAL SERVICES REQUEST FORM

15216

Special Handling (Circle as appropriate and explain in record 5)

Hazardous material
 Sample
 KAFB060808-2
 Station Name

Field ID
 USGS/WRD/NEW MEX
 Field Office

KIRTLAND AFB
 RP-SWMU'S
 Project

Site Type (circle one)

SW - Surface Water
 GW - Ground Water
 ME - Meteorological
 LK - Lake
 ES - Estuary
 SP - Spring
 SS - Special Source
 (505) 262-5341
 Phone (FTS)

File Deposition*
 Circle one)

Q - WATSTORE
 X - Lab File

Sample identification

For Laboratory Use Only

K A F B 060808-2
 Station ID or Unique Number*

463 53 600 1
 Project Account #

1991 05 31 1350 05 31 1407 N M 035 001
 Year* Month* Day* Time* Month Day Time State District/ County
 Code* User Code* Code

Analysis level codes and schedules

	6 Sample Medium**	Geologic Unit	(H) or 9 Analysis Status**	9 Analysis Source**	Hydrologic Condition**	9 Sample Type**	9 Hydrologic Event**
PARAMETER:	CHROMIUM, TOTAL		CHROMIUM, DISS		CHROMIUM HEXAVALENT TOTAL	CHROMIUM HEXAVALENT DISS.	NITRATE & NITRITE
METHOD:	SW3020/SW7191		SW3005/SW7191		SW7196	SW7196	E353.2
PARAMETER:	URANIUM, GROSS	ALPHA & BETA		VOC		VOX	TOC/TOX
METHOD:	A711B, E900			SW5030/SW8240	SW5030/SW8010		E415.1/SW9020
PARAMETER:							
METHOD:							

Chain-of-Custody Record

PROJECT NAME KIRTLAND AFB-IRP, SWMU'S PROJECT NO. 463536001 P.O. NO.

Relinquished by: (Signature) Mike Roybal	Received by: (Signature) FEDERAL EXPRESS	Date 5/31/91	Time 1630
Relinquished by: (Signature)	Received by: (Signature)	Date	Time
Relinquished by: (Signature)	Received at lab by: (Signature) Joe Hines	Date 6/1/91	Time 0800
Relinquished from lab by: (Signature)	Received by: (Signature)	Date	Time

Comments (Only 50 characters stored in NWIS)

Record 5 NW GOLF COURSE POND
 Record 6

Total number of sample bottles for this request: 7

SHIP TO:

Enseco-Rocky Mountain Analytical
 4955 Yarrow Street
 Arvada, CO 80002
 (303) 421-6611
 LINDSAY BREYER

ENSECO ANALYTICAL SERVICES REQUEST FORM

15216 03

Special Handling (Circle as appropriate and explain in record 5)

Hazardous material
EQUIPMENT BLANK
KAFB060809-2
Station Name

Field ID
USGS/WRD/NEW MEX
Field Office

KIRTLAND AFB
RP-SWMU'S
Project

SW - Surface Water
GW - Ground Water
ME - Meteorological
Collector

Site Type (circle one)

LK - Lake
ES - Estuary
SP - Spring
SS - Special Source
Phone: (505) 262-5300
FTS

File Deposition*
Circle one)
Q - WATSTORE
X - Lab File

Sample identification

For Laboratory Use Only

K A F B 060809-2
Station ID or Unique Number*

463 536001
Project Account #

9 9 1 05 31 1455 05 31 1505 N M 035 001
Year* Month* Day* Time* Month* Day* Time* State Code* District/ User Code* County Code

Analysis level codes and schedules

PARAMETER:	6 Sample Medium**	Geologic Unit	(H) or 9 Analysis Status**	9 Analysis Source**	Hydrologic Condition**	9 Sample	9 Hydrologic Event**
CHROMIUM, TOTAL	SW3020/SW7191		CHROMIUM, DISS	SW3005/SW7191	CHROMIUM HEXAVALENT TOTAL	SW7196	
METHOD:					CHROMIUM HEXAVALENT DISS.	SW7196	NITRATE & NITRITE
CHROMIUM, DISS			CHROMIUM, GROSS BETA				E353.2
METHOD:							
URANIUM, GROSS BETA	A711B, E900						TOC/TOX
METHOD:							E415.1/SW9020
PARAMETER:							
METHOD:							

Chain-of-Custody Record

PROJECT NAME KIRTLAND AFB-IRP, SWMU'S PROJECT NO. 463536001 P.O. NO.

Relinquished by: (Signature)	Received by: (Signature)	Date	Time
<i>Miho Royal</i>	FEDERAL EXPRESS	5/31/91	16:30
Relinquished by: (Signature)	Received by: (Signature)	Date	Time
Relinquished by: (Signature)	Received at lab by: (Signature)	Date	Time
	<i>[Signature]</i>	6/1/91	08:00
Relinquished from lab by: (Signature)	Received by: (Signature)	Date	Time

Comments (Only 50 characters stored in NWIS)

Record 5 NEW GOLF COURSE POND - EQUIPMENT BLANK

Record 6

Total number of sample bottles for this request: 7

SHIP TO:

Enseco-Rocky Mountain Analytical
4955 Yarrow Street
Arvada, CO 80002
(303) 421-6611
LINDSAY BREYER

ATTENTION: