



From: CLARK, SCOTT C GS-13 USAF AFCEC/CZO [mailto:scott.clark@us.af.mil]
Sent: Monday, March 20, 2017 4:01 PM
To: Hunter, Michelle, NMENV <Michelle.Hunter@state.nm.us>; Agnew, Diane, NMENV <Diane.Agnew@state.nm.us>; Kieling, John, NMENV <john.kieling@state.nm.us>
Cc: BODOUR, ADRIA A CIV USAF HAF AFCEC/CZR <adria.bodour.1@us.af.mil>; LYNNE, KATHRYN D HQE USAF AFGSC 377 MSG/SAF/IEE <kathryn.lynnes@us.af.mil>; Julie McNeill (jmcneill@portageinc.com) <jmcneill@portageinc.com>; Simpler, Trent SPA (Trent.Simpler@usace.army.mil) <Trent.Simpler@usace.army.mil>; 'Salazar, Carlos F SPA' <Carlos.F.Salazar@usace.army.mil>; Linda Dreeland <Linda.Dreeland@usace.army.mil>; Phaneuf, Mark J SPA (Mark.J.Phaneuf@usace.army.mil) <Mark.J.Phaneuf@usace.army.mil>; O'GRADY, HOLLY M GS-12 USAF AFCEC AFCEC/CZOW <holly.ograde@us.af.mil>; PINO, ANTIONETTE R CTR USAF AFCEC AFCEC/CZOW <antionette.pino.ctr@us.af.mil>; SEGURA, CHRISTOPHER G GS-13 USAF AFCEC/CZO <christopher.segura.2@us.af.mil>
Subject: 1203 Written Notification to NMED GWQB

Hi All,

As a follow-up to the initial verbal notification of a release today that was given to Diane Agnew, GWQB, please find the written notification of the release below. The drilling fluids had been analyzed prior to this release and no hazardous constituents were present, and lab data and photos are attached for reference.

Written Notification of Release pursuant to 20.6.2.1203.A NMAC: Drilling fluids leaking from bottom of KAFB-106239 lined roll off (Roll off number KAFB-106239-LC)

(a) Person In Charge:

Scott Clark
Kirtland IST, AFCEC/CZO
2050 Wyoming Boulevard Southeast
Kirtland Air Force Base, New Mexico 87117-5270
505-846-9017
DSN 246-9017
Mobile 505-385-3679
scott.clark@us.af.mil

(b) Location of Discharge: Kirtland AFB, IDW storage yard on SW corner of Ridgecrest Drive SE and Perimeter RD (southwest of BFF GWTS)

(c) Date/Time Release: 18MAR17(?); drilling fluids began leaking through the liner sometime over weekend as no leak was seen on Friday 17MAR17 inspection. Leak was discovered during Monday waste inspections (20MAR17)

(d) Source and Cause of Discharge: Leakage through lined roll off in storage pending disposal off-site (Failed Liner)



(e) Description of Discharge: Drilling fluids from extraction well KAFB-106239; fluids were already analyzed for waste disposal (total metals, 8260 VOCs, 8270 SVOCs, and RCI) and contain no hazardous constituents. Sole detection of chloroform at 2.7 µg/L (PQL is 2.0 ug/L) is suspected laboratory contamination.

(f) Estimated Volume: 50 gallons; approximately 3-4 ft in width by 20 ft in length (see photographs)

(g) Actions Taken to Mitigate Potential Damage

Actions taken 20MAR17: Immediately upon discovery, soil was bermed to stop spread of fluid, and absorbent placed immediately beneath the leak location to prevent additional migration. ACT contacted (roll off leased from firm) to mobilize a replacement roll off asap. Drilling mud will be transferred to new (lined) roll off today via vacuum truck and all soil in contact with drilling mud will be removed and placed in new roll off.

Planned actions: No additional actions are planned.

Thanks, and if you have any questions just let us know.
Scott

//SIGNED//

Scott Clark
Restoration Program Manager
Kirtland IST, AFCEC/CZO
505-846-9017
DSN 246-9017
Mobile 505-385-3679
scott.clark@us.af.mil



Hall Environmental Analysis Laboratory
4901 Hawkins NE
Albuquerque, NM 87109
TEL: 505-345-3975 FAX: 505-345-4107
Website: www.hallenvironmental.com

January 17, 2017

Devon Jercinovic

EA Engineering Science & Technology
320 Gold Ave SW Suite 1210
Albuquerque, NM 87102
TEL:
FAX

RE: KAFB BFF

OrderNo.: 1701260

Dear Devon Jercinovic:

Hall Environmental Analysis Laboratory received 2 sample(s) on 1/9/2017 for the analyses presented in the following report.

These were analyzed according to EPA procedures or equivalent. To access our accredited tests please go to www.hallenvironmental.com or the state specific web sites. In order to properly interpret your results it is imperative that you review this report in its entirety. See the sample checklist and/or the Chain of Custody for information regarding the sample receipt temperature and preservation. Data qualifiers or a narrative will be provided if the sample analysis or analytical quality control parameters require a flag. When necessary, data qualifiers are provided on both the sample analysis report and the QC summary report, both sections should be reviewed. All samples are reported, as received, unless otherwise indicated. Lab measurement of analytes considered field parameters that require analysis within 15 minutes of sampling such as pH and residual chlorine are qualified as being analyzed outside of the recommended holding time.

Please don't hesitate to contact HEAL for any additional information or clarifications.

ADHS Cert #AZ0682 -- NMED-DWB Cert #NM9425 -- NMED-Micro Cert #NM0190

Sincerely,

A handwritten signature in black ink, appearing to read 'Andy Freeman', is written over a white background.

Andy Freeman
Laboratory Manager
4901 Hawkins NE
Albuquerque, NM 87109

Hall Environmental Analysis Laboratory, Inc.

CLIENT: EA Engineering Science & Technology

Client Sample ID: KAFB-106239 LC

Project: KAFB BFF

Collection Date: 1/8/2017 12:10:00 PM

Lab ID: 1701260-001

Matrix: AQUEOUS

Received Date: 1/9/2017 9:25:00 AM

| Analyses | Result | PQL | Qual | Units | DF | Date Analyzed | Batch |
|--|--------|---------|------|-------|----|-----------------------|---------------------|
| EPA METHOD 7470: MERCURY | | | | | | | Analyst: MED |
| Mercury | ND | 0.00040 | | mg/L | 1 | 1/10/2017 12:35:16 PM | 29607 |
| EPA 6010B: TOTAL RECOVERABLE METALS | | | | | | | Analyst: pmf |
| Arsenic | ND | 0.10 | | mg/L | 1 | 1/10/2017 6:24:12 PM | 29596 |
| Barium | 2.9 | 0.10 | | mg/L | 1 | 1/10/2017 6:24:12 PM | 29596 |
| Cadmium | ND | 0.010 | | mg/L | 1 | 1/10/2017 6:24:12 PM | 29596 |
| Chromium | 0.16 | 0.030 | | mg/L | 1 | 1/10/2017 6:24:12 PM | 29596 |
| Lead | 0.033 | 0.025 | | mg/L | 1 | 1/10/2017 6:24:12 PM | 29596 |
| Selenium | ND | 0.25 | | mg/L | 1 | 1/10/2017 6:24:12 PM | 29596 |
| Silver | ND | 0.025 | | mg/L | 1 | 1/10/2017 6:24:12 PM | 29596 |
| EPA METHOD 8270C: SEMIVOLATILES | | | | | | | Analyst: DAM |
| Acenaphthene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Acenaphthylene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Aniline | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Anthracene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Azobenzene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Benz(a)anthracene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Benzo(a)pyrene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Benzo(b)fluoranthene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Benzo(g,h,i)perylene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Benzo(k)fluoranthene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Benzoic acid | ND | 200 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Benzyl alcohol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Bis(2-chloroethoxy)methane | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Bis(2-chloroethyl)ether | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Bis(2-chloroisopropyl)ether | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Bis(2-ethylhexyl)phthalate | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 4-Bromophenyl phenyl ether | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Butyl benzyl phthalate | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Carbazole | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 4-Chloro-3-methylphenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 4-Chloroaniline | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2-Chloronaphthalene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2-Chlorophenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 4-Chlorophenyl phenyl ether | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Chrysene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Di-n-butyl phthalate | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Di-n-octyl phthalate | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Dibenz(a,h)anthracene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |

Refer to the QC Summary report and sample login checklist for flagged QC data and preservation information.

| | | | | |
|--------------------|----|---|----|---|
| Qualifiers: | * | Value exceeds Maximum Contaminant Level. | B | Analyte detected in the associated Method Blank |
| | D | Sample Diluted Due to Matrix | E | Value above quantitation range |
| | H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| | ND | Not Detected at the Reporting Limit | P | Sample pH Not In Range |
| | R | RPD outside accepted recovery limits | RL | Reporting Detection Limit |
| | S | % Recovery outside of range due to dilution or matrix | W | Sample container temperature is out of limit as specified |

Hall Environmental Analysis Laboratory, Inc.

CLIENT: EA Engineering Science & Technology

Client Sample ID: KAFB-106239 LC

Project: KAFB BFF

Collection Date: 1/8/2017 12:10:00 PM

Lab ID: 1701260-001

Matrix: AQUEOUS

Received Date: 1/9/2017 9:25:00 AM

| Analyses | Result | PQL | Qual | Units | DF | Date Analyzed | Batch |
|--|--------|-----|------|-------|----|----------------------|---------------------|
| EPA METHOD 8270C: SEMIVOLATILES | | | | | | | Analyst: DAM |
| Dibenzofuran | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 1,2-Dichlorobenzene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 1,3-Dichlorobenzene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 1,4-Dichlorobenzene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 3,3'-Dichlorobenzidine | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Diethyl phthalate | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Dimethyl phthalate | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2,4-Dichlorophenol | ND | 200 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2,4-Dimethylphenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 4,6-Dinitro-2-methylphenol | ND | 200 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2,4-Dinitrophenol | ND | 200 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2,4-Dinitrotoluene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2,6-Dinitrotoluene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Fluoranthene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Fluorene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Hexachlorobenzene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Hexachlorobutadiene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Hexachlorocyclopentadiene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Hexachloroethane | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Indeno(1,2,3-cd)pyrene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Isophorone | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 1-Methylnaphthalene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2-Methylnaphthalene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2-Methylphenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 3+4-Methylphenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| N-Nitrosodi-n-propylamine | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| N-Nitrosodimethylamine | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| N-Nitrosodiphenylamine | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Naphthalene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2-Nitroaniline | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 3-Nitroaniline | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 4-Nitroaniline | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Nitrobenzene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2-Nitrophenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 4-Nitrophenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Pentachlorophenol | ND | 200 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Phenanthrene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Phenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Pyrene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |

Refer to the QC Summary report and sample login checklist for flagged QC data and preservation information.

| | | | | |
|--------------------|----|---|----|---|
| Qualifiers: | * | Value exceeds Maximum Contaminant Level. | B | Analyte detected in the associated Method Blank |
| | D | Sample Diluted Due to Matrix | E | Value above quantitation range |
| | H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| | ND | Not Detected at the Reporting Limit | P | Sample pH Not In Range |
| | R | RPD outside accepted recovery limits | RL | Reporting Detection Limit |
| | S | % Recovery outside of range due to dilution or matrix | W | Sample container temperature is out of limit as specified |

Hall Environmental Analysis Laboratory, Inc.

CLIENT: EA Engineering Science & Technology

Client Sample ID: KAFB-106239 LC

Project: KAFB BFF

Collection Date: 1/8/2017 12:10:00 PM

Lab ID: 1701260-001

Matrix: AQUEOUS

Received Date: 1/9/2017 9:25:00 AM

| Analyses | Result | PQL | Qual | Units | DF | Date Analyzed | Batch |
|--|--------|-----------|------|-------|----|----------------------|---------------------|
| EPA METHOD 8270C: SEMIVOLATILES | | | | | | | Analyst: DAM |
| Pyridine | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 1,2,4-Trichlorobenzene | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2,4,5-Trichlorophenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| 2,4,6-Trichlorophenol | ND | 100 | | µg/L | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Surr: 2-Fluorophenol | 32.8 | 15-98.1 | | %Rec | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Surr: Phenol-d5 | 28.5 | 15-80.7 | | %Rec | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Surr: 2,4,6-Tribromophenol | 55.0 | 15-112 | | %Rec | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Surr: Nitrobenzene-d5 | 56.0 | 27.2-90.7 | | %Rec | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Surr: 2-Fluorobiphenyl | 40.1 | 23.3-85.6 | | %Rec | 1 | 1/13/2017 8:25:21 PM | 29601 |
| Surr: 4-Terphenyl-d14 | 26.9 | 27.6-107 | S | %Rec | 1 | 1/13/2017 8:25:21 PM | 29601 |
| EPA METHOD 8260B: VOLATILES | | | | | | | Analyst: DJF |
| Benzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Toluene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Ethylbenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Methyl tert-butyl ether (MTBE) | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2,4-Trimethylbenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,3,5-Trimethylbenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2-Dichloroethane (EDC) | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2-Dibromoethane (EDB) | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Naphthalene | ND | 4.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1-Methylnaphthalene | ND | 8.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 2-Methylnaphthalene | ND | 8.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Acetone | ND | 20 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Bromobenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Bromodichloromethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Bromoform | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Bromomethane | ND | 6.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 2-Butanone | ND | 20 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Carbon disulfide | ND | 20 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Carbon Tetrachloride | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Chlorobenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Chloroethane | ND | 4.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Chloroform | 2.7 | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Chloromethane | ND | 6.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 2-Chlorotoluene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 4-Chlorotoluene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| cis-1,2-DCE | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| cis-1,3-Dichloropropene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2-Dibromo-3-chloropropane | ND | 4.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |

Refer to the QC Summary report and sample login checklist for flagged QC data and preservation information.

| | | |
|--------------------|---|---|
| Qualifiers: | * Value exceeds Maximum Contaminant Level. | B Analyte detected in the associated Method Blank |
| | D Sample Diluted Due to Matrix | E Value above quantitation range |
| | H Holding times for preparation or analysis exceeded | J Analyte detected below quantitation limits |
| | ND Not Detected at the Reporting Limit | P Sample pH Not In Range |
| | R RPD outside accepted recovery limits | RL Reporting Detection Limit |
| | S % Recovery outside of range due to dilution or matrix | W Sample container temperature is out of limit as specified |

Hall Environmental Analysis Laboratory, Inc.

CLIENT: EA Engineering Science & Technology

Client Sample ID: KAFB-106239 LC

Project: KAFB BFF

Collection Date: 1/8/2017 12:10:00 PM

Lab ID: 1701260-001

Matrix: AQUEOUS

Received Date: 1/9/2017 9:25:00 AM

| Analyses | Result | PQL | Qual | Units | DF | Date Analyzed | Batch |
|------------------------------------|--------|--------|------|-------|----|---------------------|--------------|
| EPA METHOD 8260B: VOLATILES | | | | | | | Analyst: DJF |
| Dibromochloromethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Dibromomethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2-Dichlorobenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,3-Dichlorobenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,4-Dichlorobenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Dichlorodifluoromethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,1-Dichloroethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,1-Dichloroethene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2-Dichloropropane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,3-Dichloropropane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 2,2-Dichloropropane | ND | 4.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,1-Dichloropropene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Hexachlorobutadiene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 2-Hexanone | ND | 20 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Isopropylbenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 4-Isopropyltoluene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 4-Methyl-2-pentanone | ND | 20 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Methylene Chloride | ND | 6.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| n-Butylbenzene | ND | 6.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| n-Propylbenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| sec-Butylbenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Styrene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| tert-Butylbenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,1,1,2-Tetrachloroethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,1,1,2-Tetrachloroethane | ND | 4.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Tetrachloroethene (PCE) | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| trans-1,2-DCE | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| trans-1,3-Dichloropropene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2,3-Trichlorobenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2,4-Trichlorobenzene | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,1,1-Trichloroethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,1,2-Trichloroethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Trichloroethene (TCE) | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Trichlorofluoromethane | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| 1,2,3-Trichloropropane | ND | 4.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Vinyl chloride | ND | 2.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Xylenes, Total | ND | 3.0 | D | µg/L | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Surr: 1,2-Dichloroethane-d4 | 101 | 70-130 | D | %Rec | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Surr: 4-Bromofluorobenzene | 86.0 | 70-130 | D | %Rec | 2 | 1/9/2017 8:18:32 PM | W39912 |

Refer to the QC Summary report and sample login checklist for flagged QC data and preservation information.

| | | | | |
|--------------------|----|---|----|---|
| Qualifiers: | * | Value exceeds Maximum Contaminant Level. | B | Analyte detected in the associated Method Blank |
| | D | Sample Diluted Due to Matrix | E | Value above quantitation range |
| | H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| | ND | Not Detected at the Reporting Limit | P | Sample pH Not In Range |
| | R | RPD outside accepted recovery limits | RL | Reporting Detection Limit |
| | S | % Recovery outside of range due to dilution or matrix | W | Sample container temperature is out of limit as specified |

Hall Environmental Analysis Laboratory, Inc.

Analytical Report

Lab Order 1701260

Date Reported: 1/17/2017

CLIENT: EA Engineering Science & Technology

Client Sample ID: KAFB-106239 LC

Project: KAFB BFF

Collection Date: 1/8/2017 12:10:00 PM

Lab ID: 1701260-001

Matrix: AQUEOUS

Received Date: 1/9/2017 9:25:00 AM

| Analyses | Result | PQL | Qual | Units | DF | Date Analyzed | Batch |
|------------------------------------|--------|--------|------|-------|----|---------------------|--------------|
| EPA METHOD 8260B: VOLATILES | | | | | | | Analyst: DJF |
| Surr: Dibromofluoromethane | 113 | 70-130 | D | %Rec | 2 | 1/9/2017 8:18:32 PM | W39912 |
| Surr: Toluene-d8 | 90.8 | 70-130 | D | %Rec | 2 | 1/9/2017 8:18:32 PM | W39912 |

Refer to the QC Summary report and sample login checklist for flagged QC data and preservation information.

| | | | | |
|--------------------|----|---|----|---|
| Qualifiers: | * | Value exceeds Maximum Contaminant Level. | B | Analyte detected in the associated Method Blank |
| | D | Sample Diluted Due to Matrix | E | Value above quantitation range |
| | H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| | ND | Not Detected at the Reporting Limit | P | Sample pH Not In Range |
| | R | RPD outside accepted recovery limits | RL | Reporting Detection Limit |
| | S | % Recovery outside of range due to dilution or matrix | W | Sample container temperature is out of limit as specified |

Hall Environmental Analysis Laboratory, Inc.

CLIENT: EA Engineering Science & Technology

Client Sample ID: Trip Blank

Project: KAFB BFF

Collection Date:

Lab ID: 1701260-002

Matrix: AQUEOUS

Received Date: 1/9/2017 9:25:00 AM

| Analyses | Result | PQL | Qual | Units | DF | Date Analyzed | Batch |
|------------------------------------|--------|-----|------|-------|----|---------------------|--------------|
| EPA METHOD 8260B: VOLATILES | | | | | | | Analyst: DJF |
| Benzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Toluene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Ethylbenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Methyl tert-butyl ether (MTBE) | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2,4-Trimethylbenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,3,5-Trimethylbenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2-Dichloroethane (EDC) | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2-Dibromoethane (EDB) | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Naphthalene | ND | 2.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1-Methylnaphthalene | ND | 4.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 2-Methylnaphthalene | ND | 4.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Acetone | ND | 10 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Bromobenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Bromodichloromethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Bromoform | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Bromomethane | ND | 3.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 2-Butanone | ND | 10 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Carbon disulfide | ND | 10 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Carbon Tetrachloride | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Chlorobenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Chloroethane | ND | 2.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Chloroform | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Chloromethane | ND | 3.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 2-Chlorotoluene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 4-Chlorotoluene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| cis-1,2-DCE | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| cis-1,3-Dichloropropene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2-Dibromo-3-chloropropane | ND | 2.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Dibromochloromethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Dibromomethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2-Dichlorobenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,3-Dichlorobenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,4-Dichlorobenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Dichlorodifluoromethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,1-Dichloroethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,1-Dichloroethene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2-Dichloropropane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,3-Dichloropropane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 2,2-Dichloropropane | ND | 2.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |

Refer to the QC Summary report and sample login checklist for flagged QC data and preservation information.

| | | | | |
|--------------------|----|---|----|---|
| Qualifiers: | * | Value exceeds Maximum Contaminant Level. | B | Analyte detected in the associated Method Blank |
| | D | Sample Diluted Due to Matrix | E | Value above quantitation range |
| | H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| | ND | Not Detected at the Reporting Limit | P | Sample pH Not In Range |
| | R | RPD outside accepted recovery limits | RL | Reporting Detection Limit |
| | S | % Recovery outside of range due to dilution or matrix | W | Sample container temperature is out of limit as specified |

Hall Environmental Analysis Laboratory, Inc.

CLIENT: EA Engineering Science & Technology

Client Sample ID: Trip Blank

Project: KAFB BFF

Collection Date:

Lab ID: 1701260-002

Matrix: AQUEOUS

Received Date: 1/9/2017 9:25:00 AM

| Analyses | Result | PQL | Qual | Units | DF | Date Analyzed | Batch |
|------------------------------------|--------|--------|------|-------|----|---------------------|--------------|
| EPA METHOD 8260B: VOLATILES | | | | | | | Analyst: DJF |
| 1,1-Dichloropropene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Hexachlorobutadiene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 2-Hexanone | ND | 10 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Isopropylbenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 4-Isopropyltoluene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 4-Methyl-2-pentanone | ND | 10 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Methylene Chloride | ND | 3.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| n-Butylbenzene | ND | 3.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| n-Propylbenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| sec-Butylbenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Styrene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| tert-Butylbenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,1,1,2-Tetrachloroethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,1,2,2-Tetrachloroethane | ND | 2.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Tetrachloroethene (PCE) | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| trans-1,2-DCE | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| trans-1,3-Dichloropropene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2,3-Trichlorobenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2,4-Trichlorobenzene | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,1,1-Trichloroethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,1,2-Trichloroethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Trichloroethene (TCE) | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Trichlorofluoromethane | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| 1,2,3-Trichloropropane | ND | 2.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Vinyl chloride | ND | 1.0 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Xylenes, Total | ND | 1.5 | | µg/L | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Surr: 1,2-Dichloroethane-d4 | 98.4 | 70-130 | | %Rec | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Surr: 4-Bromofluorobenzene | 86.2 | 70-130 | | %Rec | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Surr: Dibromofluoromethane | 115 | 70-130 | | %Rec | 1 | 1/9/2017 8:48:00 PM | W39912 |
| Surr: Toluene-d8 | 90.0 | 70-130 | | %Rec | 1 | 1/9/2017 8:48:00 PM | W39912 |

Refer to the QC Summary report and sample login checklist for flagged QC data and preservation information.

| | | | | |
|--------------------|----|---|----|---|
| Qualifiers: | * | Value exceeds Maximum Contaminant Level. | B | Analyte detected in the associated Method Blank |
| | D | Sample Diluted Due to Matrix | E | Value above quantitation range |
| | H | Holding times for preparation or analysis exceeded | J | Analyte detected below quantitation limits |
| | ND | Not Detected at the Reporting Limit | P | Sample pH Not In Range |
| | R | RPD outside accepted recovery limits | RL | Reporting Detection Limit |
| | S | % Recovery outside of range due to dilution or matrix | W | Sample container temperature is out of limit as specified |



Wet Chemistry by Method 9012 B

| Analyte | Result | Qualifier | RDL | Dilution | Analysis date / time | Batch |
|------------------|------------|-----------|-----------------|----------|----------------------|----------|
| Reactive Cyanide | mg/l ND | | mg/l 0.00500 | 1 | 01/11/2017 00:53 | WG942128 |

²Tc

Wet Chemistry by Method 9034-9030B

| Analyte | Result | Qualifier | RDL | Dilution | Analysis date / time | Batch |
|------------------|--------------|-----------|----------------|----------|----------------------|----------|
| Reactive Sulfide | mg/l 1.25 | | mg/l 0.0500 | 1 | 01/10/2017 18:57 | WG942096 |

³Ss

⁴Cn

Wet Chemistry by Method 9040C

| Analyte | Result | Qualifier | Dilution | Analysis date / time | Batch |
|-------------------|------------|-----------|----------|----------------------|----------|
| Corrosivity by pH | su 9.90 | | 1 | 01/12/2017 10:23 | WG942433 |

⁵Sr

⁶Qc

Sample Narrative:

9040C L882873-01 WG942433: 9.90 at 17.7c

⁷Gl

⁸Al

Wet Chemistry by Method D93/1010A

| Analyte | Result | Qualifier | Dilution | Analysis date / time | Batch |
|------------|-----------------------|-----------|----------|----------------------|----------|
| Flashpoint | deg F DNF at 170 F | | 1 | 01/11/2017 01:19 | WG942008 |

⁹Sc

WG942128

QUALITY CONTROL SUMMARY

ONE LAB. NATIONWIDE.



Wet Chemistry by Method 9012 B

L882873-01

Method Blank (MB)

(MB) R3189873-1 01/11/17 00:23

| Analyte | MB Result | MB Qualifier | MB MDL | MB RDL |
|------------------|-----------|--------------|--------|---------|
| Reactive Cyanide | mg/l | U | 0.0018 | 0.00500 |

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) R3189873-2 01/11/17 00:24 • (LCSD) R3189873-3 01/11/17 00:25

| Analyte | Spike Amount | LCS Result | LCSD Result | LCS Rec. | LCSD Rec. | Rec. Limits | LCS Qualifier | LCSD Qualifier | RPD | RPD Limits |
|------------------|--------------|------------|-------------|----------|-----------|-------------|---------------|----------------|-----|------------|
| Reactive Cyanide | mg/l | mg/l | mg/l | % | % | % | | | % | % |
| | 0.100 | 0.110 | 0.105 | 110 | 105 | 85-115 | | | 5 | 20 |

1

2 Ss

3 Cn

4 Sr

5 Qc

6 GI

7

8 Sc

ACCOUNT:

Halt Environmental Analysis Laboratory

PROJECT:

SDG:

L882873

DATE/TIME:

01/12/17 13:50

WG942096

Wet Chemistry by Method 9034-9030B

QUALITY CONTROL SUMMARY

L882873-01

ONE LAB. NATIONWIDE



Method Blank (MB)

(MB) R3189848-1 01/10/17 18:50

| Analyte | MB Result mg/l | MB Qualifier | MB MDL mg/l | MB RDL mg/l |
|------------------|-------------------|--------------|----------------|----------------|
| Reactive Sulfide | U | | 0.00650 | 0.0500 |

Laboratory Control Sample (LCS) - Laboratory Control Sample Duplicate (LCSD)

(LCS) R3189848-2 01/10/17 18:51 • (LCSD) R3189848-3 01/10/17 18:51

| Analyte | Spike Amount mg/l | LCS Result mg/l | LCSD Result mg/l | LCS Rec. % | LCSD Rec. % | Rec. Limits % | LCS Qualifier | LCSD Qualifier | RPD % | RPD Limits % |
|------------------|----------------------|--------------------|---------------------|---------------|----------------|------------------|---------------|----------------|----------|-----------------|
| Reactive Sulfide | 0.500 | 0.548 | 0.549 | 110 | 110 | 85.0-115 | | | 0.000 | 20 |

1 Tl

3 Ss

4 Cn

5 Sr

6 Qc

7 Gi

8 Pb

9 Sc

ACCOUNT:

Hall Environmental Analysis Laboratory

PROJECT:

SDG:

L882873

DATE/TIME:

01/12/17 13:50

WG942433

Wet Chemistry by Method 9040C

QUALITY CONTROL SUMMARY

L882873-01

ONE LAB. NATIONWIDE.



L882301-01 Original Sample (OS) • Duplicate (DUP)

(OS) L882301-01 01/12/17 10:23 • (DUP) WG942433-3 01/12/17 10:23

| Analyte | Original Result | DUP Result | Dilution | DUP RPD | DUP Qualifier | DUP RPD Limits |
|-------------------|-----------------|------------|----------|---------|---------------|----------------|
| Corrosivity by pH | su | su | 1 | % | | % |
| | 4.07 | 4.08 | 1 | 0.245 | | 1 |

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) WG942433-1 01/12/17 10:23 • (LCSD) WG942433-2 01/12/17 10:23

| Analyte | Spike Amount | LCS Result | LCSD Result | LCS Rec. | LCSD Rec. | Rec. Limits | LCS Qualifier | LCSD Qualifier | RPD | RPD Limits |
|-------------------|--------------|------------|-------------|----------|-----------|-------------|---------------|----------------|-------|------------|
| Corrosivity by pH | su | su | su | % | % | % | | | % | % |
| | 6.07 | 6.09 | 6.07 | 100 | 100 | 98.4-102 | | | 0.329 | 1 |

1

2 Ss

4 Cn

5 Sr

6 Qc

7 Gi

8

9 Sc

ACCOUNT:

Hall Environmental Analysis Laboratory

PROJECT:

SDG:

L882873

DATE/TIME:

01/12/17 13:50

WG942008

Wet Chemistry by Method D93/1010A

QUALITY CONTROL SUMMARY

L882873-01

ONE LAB. NATIONWIDE



L882979-07 Original Sample (OS) • Duplicate (DUP)

(OS) L882979-07 01/11/17 01:19 • (DUP) WG942008-3 01/11/17 01:19

| Analyte | Original Result | DUP Result | Dilution | DUP RPD | DUP Qualifier | DUP RPD Limits |
|------------|-----------------|--------------|----------|---------|---------------|----------------|
| | deg F | deg F | | % | | % |
| Flashpoint | DNF at 170 F | DNF at 170 F | 1 | 0.000 | | 20 |

Laboratory Control Sample (LCS) • Laboratory Control Sample Duplicate (LCSD)

(LCS) WG942008-1 01/11/17 01:19 • (LCSD) WG942008-2 01/11/17 01:19

| Analyte | Spike Amount | LCS Result | LCSD Result | LCS Rec. | LCSD Rec. | Rec. Limits | LCS Qualifier | LCSD Qualifier | RPD | RPD Limits |
|------------|--------------|------------|-------------|----------|-----------|-------------|---------------|----------------|-------|------------|
| | deg F | deg F | deg F | % | % | % | | | % | % |
| Flashpoint | 82.0 | 82.7 | 82.7 | 101 | 101 | 96.0-104 | | | 0.000 | 7 |



ACCOUNT:

Hall Environmental Analysis Laboratory

PROJECT:

SDG:

L882873

DATE/TIME:

01/12/17 13:50

GLOSSARY OF TERMS

ONE LAB. NATIONWIDE.



Abbreviations and Definitions

| | |
|-----------------|---|
| SDG | Sample Delivery Group. |
| MDL | Method Detection Limit. |
| RDL | Reported Detection Limit. |
| ND | Not detected at the Reporting Limit (or MDL where applicable). |
| U | Not detected at the Reporting Limit (or MDL where applicable). |
| RPD | Relative Percent Difference. |
| Original Sample | The non-spiked sample in the prep batch used to determine the Relative Percent Difference (RPD) from a quality control sample. The Original Sample may not be included within the reported SDG. |
| Rec. | Recovery. |

| Qualifier | Description |
|-----------|-------------|
|-----------|-------------|

The remainder of this page intentionally left blank, there are no qualifiers applied to this SDG.

²Tc

³Ss

⁴Cn

⁵Sr

⁶Qc

⁷Gl

⁸Al

⁹Sc

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology

Project: KAFB BFF

| | | | |
|------------|------------|--------------------------------|--|
| Sample ID | rb | SampType: MBLK | TestCode: EPA Method 8260B: VOLATILES |
| Client ID: | PBW | Batch ID: W39912 | RunNo: 39912 |
| Prep Date: | | Analysis Date: 1/9/2017 | SeqNo: 1250932 Units: µg/L |

| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
|--------------------------------|--------|-----|-----------|-------------|------|----------|-----------|------|----------|------|
| Benzene | ND | 1.0 | | | | | | | | |
| Toluene | ND | 1.0 | | | | | | | | |
| Ethylbenzene | ND | 1.0 | | | | | | | | |
| Methyl tert-butyl ether (MTBE) | ND | 1.0 | | | | | | | | |
| 1,2,4-Trimethylbenzene | ND | 1.0 | | | | | | | | |
| 1,3,5-Trimethylbenzene | ND | 1.0 | | | | | | | | |
| 1,2-Dichloroethane (EDC) | ND | 1.0 | | | | | | | | |
| 1,2-Dibromoethane (EDB) | ND | 1.0 | | | | | | | | |
| Naphthalene | ND | 2.0 | | | | | | | | |
| 1-Methylnaphthalene | ND | 4.0 | | | | | | | | |
| 2-Methylnaphthalene | ND | 4.0 | | | | | | | | |
| Acetone | ND | 10 | | | | | | | | |
| Bromobenzene | ND | 1.0 | | | | | | | | |
| Bromodichloromethane | ND | 1.0 | | | | | | | | |
| Bromoform | ND | 1.0 | | | | | | | | |
| Bromomethane | ND | 3.0 | | | | | | | | |
| 2-Butanone | ND | 10 | | | | | | | | |
| Carbon disulfide | ND | 10 | | | | | | | | |
| Carbon Tetrachloride | ND | 1.0 | | | | | | | | |
| Chlorobenzene | ND | 1.0 | | | | | | | | |
| Chloroethane | ND | 2.0 | | | | | | | | |
| Chloroform | ND | 1.0 | | | | | | | | |
| Chloromethane | ND | 3.0 | | | | | | | | |
| 2-Chlorotoluene | ND | 1.0 | | | | | | | | |
| 4-Chlorotoluene | ND | 1.0 | | | | | | | | |
| cis-1,2-DCE | ND | 1.0 | | | | | | | | |
| cis-1,3-Dichloropropene | ND | 1.0 | | | | | | | | |
| 1,2-Dibromo-3-chloropropane | ND | 2.0 | | | | | | | | |
| Dibromochloromethane | ND | 1.0 | | | | | | | | |
| Dibromomethane | ND | 1.0 | | | | | | | | |
| 1,2-Dichlorobenzene | ND | 1.0 | | | | | | | | |
| 1,3-Dichlorobenzene | ND | 1.0 | | | | | | | | |
| 1,4-Dichlorobenzene | ND | 1.0 | | | | | | | | |
| Dichlorodifluoromethane | ND | 1.0 | | | | | | | | |
| 1,1-Dichloroethane | ND | 1.0 | | | | | | | | |
| 1,1-Dichloroethene | ND | 1.0 | | | | | | | | |
| 1,2-Dichloropropane | ND | 1.0 | | | | | | | | |
| 1,3-Dichloropropane | ND | 1.0 | | | | | | | | |
| 2,2-Dichloropropane | ND | 2.0 | | | | | | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology
Project: KAFB BFF

| | | | | | | | | | | |
|------------|-----|----------------|----------|-----------|-----------------------------|--------|------|--|--|--|
| Sample ID | rb | SampType: | MBLK | TestCode: | EPA Method 8260B: VOLATILES | | | | | |
| Client ID: | PBW | Batch ID: | W39912 | RunNo: | 39912 | | | | | |
| Prep Date: | | Analysis Date: | 1/9/2017 | SeqNo: | 1250932 | Units: | µg/L | | | |

| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
|-----------------------------|--------|-----|-----------|-------------|------|----------|-----------|------|----------|------|
| 1,1-Dichloropropene | ND | 1.0 | | | | | | | | |
| Hexachlorobutadiene | ND | 1.0 | | | | | | | | |
| 2-Hexanone | ND | 10 | | | | | | | | |
| Isopropylbenzene | ND | 1.0 | | | | | | | | |
| 4-Isopropyltoluene | ND | 1.0 | | | | | | | | |
| 4-Methyl-2-pentanone | ND | 10 | | | | | | | | |
| Methylene Chloride | ND | 3.0 | | | | | | | | |
| n-Butylbenzene | ND | 3.0 | | | | | | | | |
| n-Propylbenzene | ND | 1.0 | | | | | | | | |
| sec-Butylbenzene | ND | 1.0 | | | | | | | | |
| Styrene | ND | 1.0 | | | | | | | | |
| tert-Butylbenzene | ND | 1.0 | | | | | | | | |
| 1,1,1,2-Tetrachloroethane | ND | 1.0 | | | | | | | | |
| 1,1,2,2-Tetrachloroethane | ND | 2.0 | | | | | | | | |
| Tetrachloroethane (PCE) | ND | 1.0 | | | | | | | | |
| trans-1,2-DCE | ND | 1.0 | | | | | | | | |
| trans-1,3-Dichloropropene | ND | 1.0 | | | | | | | | |
| 1,2,3-Trichlorobenzene | ND | 1.0 | | | | | | | | |
| 1,2,4-Trichlorobenzene | ND | 1.0 | | | | | | | | |
| 1,1,1-Trichloroethane | ND | 1.0 | | | | | | | | |
| 1,1,2-Trichloroethane | ND | 1.0 | | | | | | | | |
| Trichloroethene (TCE) | ND | 1.0 | | | | | | | | |
| Trichlorofluoromethane | ND | 1.0 | | | | | | | | |
| 1,2,3-Trichloropropane | ND | 2.0 | | | | | | | | |
| Vinyl chloride | ND | 1.0 | | | | | | | | |
| Xylenes, Total | ND | 1.5 | | | | | | | | |
| Surr: 1,2-Dichloroethane-d4 | 9.1 | | 10.00 | | 91.4 | 70 | 130 | | | |
| Surr: 4-Bromofluorobenzene | 9.1 | | 10.00 | | 91.2 | 70 | 130 | | | |
| Surr: Dibromofluoromethane | 9.7 | | 10.00 | | 97.5 | 70 | 130 | | | |
| Surr: Toluene-d8 | 8.9 | | 10.00 | | 88.9 | 70 | 130 | | | |

| | | | | | | | | | | |
|------------|-----------|----------------|----------|-----------|-----------------------------|--------|------|--|--|--|
| Sample ID | 100ng lcs | SampType: | LCS | TestCode: | EPA Method 8260B: VOLATILES | | | | | |
| Client ID: | LCSW | Batch ID: | W39912 | RunNo: | 39912 | | | | | |
| Prep Date: | | Analysis Date: | 1/9/2017 | SeqNo: | 1250933 | Units: | µg/L | | | |

| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
|---------------|--------|-----|-----------|-------------|------|----------|-----------|------|----------|------|
| Benzene | 23 | 1.0 | 20.00 | 0 | 115 | 70 | 130 | | | |
| Toluene | 19 | 1.0 | 20.00 | 0 | 97.2 | 70 | 130 | | | |
| Chlorobenzene | 19 | 1.0 | 20.00 | 0 | 96.9 | 70 | 130 | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology

Project: KAFB BFF

| Sample ID | 100ng lcs | SampType: | LCS | TestCode: | EPA Method 8260B: VOLATILES | | | | | |
|-----------------------------|-----------|----------------|-----------|-------------|-----------------------------|----------|-----------|------|----------|------|
| Client ID: | LCSW | Batch ID: | W39912 | RunNo: | 39912 | | | | | |
| Prep Date: | | Analysis Date: | 1/9/2017 | SeqNo: | 1250933 | Units: | µg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| 1,1-Dichloroethene | 20 | 1.0 | 20.00 | 0 | 102 | 70 | 130 | | | |
| Trichloroethene (TCE) | 22 | 1.0 | 20.00 | 0 | 109 | 70 | 130 | | | |
| Surr: 1,2-Dichloroethane-d4 | 9.4 | | 10.00 | | 93.8 | 70 | 130 | | | |
| Surr: 4-Bromofluorobenzene | 9.0 | | 10.00 | | 89.5 | 70 | 130 | | | |
| Surr: Dibromofluoromethane | 9.8 | | 10.00 | | 98.4 | 70 | 130 | | | |
| Surr: Toluene-d8 | 9.1 | | 10.00 | | 91.5 | 70 | 130 | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology
Project: KAFB BFF

| Sample ID | SampType: LCS | | TestCode: EPA Method 8270C: Semivolatiles | | | | | | | |
|----------------------------|--------------------------|-----|---|-------------|------|----------|-----------|------|----------|------|
| Client ID: | Batch ID: 29601 | | RunNo: 40049 | | | | | | | |
| Prep Date: 1/9/2017 | Analysis Date: 1/13/2017 | | SeqNo: 1255086 Units: µg/L | | | | | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Acenaphthene | 60 | 10 | 100.0 | 0 | 60.1 | 42.9 | 100 | | | |
| 4-Chloro-3-methylphenol | 140 | 10 | 200.0 | 0 | 68.6 | 36.2 | 110 | | | |
| 2-Chlorophenol | 130 | 10 | 200.0 | 0 | 65.3 | 33.4 | 97.8 | | | |
| 1,4-Dichlorobenzene | 57 | 10 | 100.0 | 0 | 56.7 | 32.8 | 79.3 | | | |
| 2,4-Dinitrotoluene | 53 | 10 | 100.0 | 0 | 53.3 | 34.9 | 107 | | | |
| N-Nitrosodi-n-propylamine | 70 | 10 | 100.0 | 0 | 70.2 | 30.7 | 111 | | | |
| 4-Nitrophenol | 69 | 10 | 200.0 | 0 | 34.6 | 15 | 91.9 | | | |
| Pentachlorophenol | 110 | 20 | 200.0 | 0 | 54.9 | 33.3 | 93.5 | | | |
| Phenol | 75 | 10 | 200.0 | 0 | 37.5 | 20.9 | 86.4 | | | |
| Pyrene | 60 | 10 | 100.0 | 0 | 60.2 | 45.6 | 111 | | | |
| 1,2,4-Trichlorobenzene | 63 | 10 | 100.0 | 0 | 63.3 | 38.7 | 88.2 | | | |
| Surr: 2-Fluorophenol | 96 | | 200.0 | | 48.2 | 15 | 98.1 | | | |
| Surr: Phenol-d5 | 80 | | 200.0 | | 40.0 | 15 | 80.7 | | | |
| Surr: 2,4,6-Tribromophenol | 140 | | 200.0 | | 70.8 | 15 | 112 | | | |
| Surr: Nitrobenzene-d5 | 68 | | 100.0 | | 67.7 | 27.2 | 90.7 | | | |
| Surr: 2-Fluorobiphenyl | 63 | | 100.0 | | 62.7 | 23.3 | 85.6 | | | |
| Surr: 4-Terphenyl-d14 | 62 | | 100.0 | | 61.7 | 27.6 | 107 | | | |

| Sample ID | SampType: LCS | | TestCode: EPA Method 8270C: Semivolatiles | | | | | | | |
|----------------------------|--------------------------|-----|---|-------------|------|----------|-----------|------|----------|------|
| Client ID: | Batch ID: 29601 | | RunNo: 40049 | | | | | | | |
| Prep Date: 1/9/2017 | Analysis Date: 1/13/2017 | | SeqNo: 1255088 Units: µg/L | | | | | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Acenaphthene | 49 | 10 | 100.0 | 0 | 48.6 | 42.9 | 100 | | | |
| 4-Chloro-3-methylphenol | 100 | 10 | 200.0 | 0 | 52.4 | 36.2 | 110 | | | |
| 2-Chlorophenol | 88 | 10 | 200.0 | 0 | 44.1 | 33.4 | 97.8 | | | |
| 1,4-Dichlorobenzene | 39 | 10 | 100.0 | 0 | 38.9 | 32.8 | 79.3 | | | |
| 2,4-Dinitrotoluene | 43 | 10 | 100.0 | 0 | 43.2 | 34.9 | 107 | | | |
| N-Nitrosodi-n-propylamine | 53 | 10 | 100.0 | 0 | 52.9 | 30.7 | 111 | | | |
| 4-Nitrophenol | 51 | 10 | 200.0 | 0 | 25.5 | 15 | 91.9 | | | |
| Pentachlorophenol | 87 | 20 | 200.0 | 0 | 43.4 | 33.3 | 93.5 | | | |
| Phenol | 50 | 10 | 200.0 | 0 | 25.2 | 20.9 | 86.4 | | | |
| Pyrene | 52 | 10 | 100.0 | 0 | 51.6 | 45.6 | 111 | | | |
| 1,2,4-Trichlorobenzene | 51 | 10 | 100.0 | 0 | 50.6 | 38.7 | 88.2 | | | |
| Surr: 2-Fluorophenol | 58 | | 200.0 | | 29.1 | 15 | 98.1 | | | |
| Surr: Phenol-d5 | 53 | | 200.0 | | 26.7 | 15 | 80.7 | | | |
| Surr: 2,4,6-Tribromophenol | 110 | | 200.0 | | 53.2 | 15 | 112 | | | |
| Surr: Nitrobenzene-d5 | 49 | | 100.0 | | 49.3 | 27.2 | 90.7 | | | |
| Surr: 2-Fluorobiphenyl | 46 | | 100.0 | | 46.1 | 23.3 | 85.6 | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology

Project: KAFB BFF

| | | | | | | | | | | |
|-----------------------|-------------------|----------------|------------------|-------------|--|----------|-------------|------|----------|------|
| Sample ID | icsd-29601 | SampType: | LCS | TestCode: | EPA Method 8270C: Semivolatiles | | | | | |
| Client ID: | LCSW | Batch ID: | 29601 | RunNo: | 40049 | | | | | |
| Prep Date: | 1/9/2017 | Analysis Date: | 1/13/2017 | SeqNo: | 1255088 | Units: | µg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Surr: 4-Terphenyl-d14 | 43 | | 100.0 | | 42.6 | 27.6 | 107 | | | |

| | | | | | | | | | | |
|------------|-----------------|----------------|------------------|-------------|--|----------|-------------|------|----------|------|
| Sample ID | mb-29601 | SampType: | MBLK | TestCode: | EPA Method 8270C: Semivolatiles | | | | | |
| Client ID: | PBW | Batch ID: | 29601 | RunNo: | 40049 | | | | | |
| Prep Date: | 1/9/2017 | Analysis Date: | 1/13/2017 | SeqNo: | 1255089 | Units: | µg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |

| | | | | | | | | | | |
|-----------------------------|----|----|--|--|--|--|--|--|--|--|
| Acenaphthene | ND | 10 | | | | | | | | |
| Acenaphthylene | ND | 10 | | | | | | | | |
| Aniline | ND | 10 | | | | | | | | |
| Anthracene | ND | 10 | | | | | | | | |
| Azobenzene | ND | 10 | | | | | | | | |
| Benz(a)anthracene | ND | 10 | | | | | | | | |
| Benzo(a)pyrene | ND | 10 | | | | | | | | |
| Benzo(b)fluoranthene | ND | 10 | | | | | | | | |
| Benzo(g,h,i)perylene | ND | 10 | | | | | | | | |
| Benzo(k)fluoranthene | ND | 10 | | | | | | | | |
| Benzoic acid | ND | 20 | | | | | | | | |
| Benzyl alcohol | ND | 10 | | | | | | | | |
| Bis(2-chloroethoxy)methane | ND | 10 | | | | | | | | |
| Bis(2-chloroethyl)ether | ND | 10 | | | | | | | | |
| Bis(2-chloroisopropyl)ether | ND | 10 | | | | | | | | |
| Bis(2-ethylhexyl)phthalate | ND | 10 | | | | | | | | |
| 4-Bromophenyl phenyl ether | ND | 10 | | | | | | | | |
| Butyl benzyl phthalate | ND | 10 | | | | | | | | |
| Carbazole | ND | 10 | | | | | | | | |
| 4-Chloro-3-methylphenol | ND | 10 | | | | | | | | |
| 4-Chloroaniline | ND | 10 | | | | | | | | |
| 2-Chloronaphthalene | ND | 10 | | | | | | | | |
| 2-Chlorophenol | ND | 10 | | | | | | | | |
| 4-Chlorophenyl phenyl ether | ND | 10 | | | | | | | | |
| Chrysene | ND | 10 | | | | | | | | |
| Di-n-butyl phthalate | ND | 10 | | | | | | | | |
| Di-n-octyl phthalate | ND | 10 | | | | | | | | |
| Dibenz(a,h)anthracene | ND | 10 | | | | | | | | |
| Dibenzofuran | ND | 10 | | | | | | | | |
| 1,2-Dichlorobenzene | ND | 10 | | | | | | | | |
| 1,3-Dichlorobenzene | ND | 10 | | | | | | | | |
| 1,4-Dichlorobenzene | ND | 10 | | | | | | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology
Project: KAFB BFF

| Sample ID | mb-29601 | SampType: | MBLK | TestCode: | EPA Method 8270C: Semivolatiles | | | | | |
|------------|-----------------|----------------|------------------|-------------|--|----------|-------------|------|----------|------|
| Client ID: | PBW | Batch ID: | 29601 | RunNo: | 40049 | | | | | |
| Prep Date: | 1/9/2017 | Analysis Date: | 1/13/2017 | SeqNo: | 1255089 | Units: | µg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |

| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
|----------------------------|--------|-----|-----------|-------------|------|----------|-----------|------|----------|------|
| 3,3'-Dichlorobenzidine | ND | 10 | | | | | | | | |
| Diethyl phthalate | ND | 10 | | | | | | | | |
| Dimethyl phthalate | ND | 10 | | | | | | | | |
| 2,4-Dichlorophenol | ND | 20 | | | | | | | | |
| 2,4-Dimethylphenol | ND | 10 | | | | | | | | |
| 4,6-Dinitro-2-methylphenol | ND | 20 | | | | | | | | |
| 2,4-Dinitrophenol | ND | 20 | | | | | | | | |
| 2,4-Dinitrotoluene | ND | 10 | | | | | | | | |
| 2,6-Dinitrotoluene | ND | 10 | | | | | | | | |
| Fluoranthene | ND | 10 | | | | | | | | |
| Fluorene | ND | 10 | | | | | | | | |
| Hexachlorobenzene | ND | 10 | | | | | | | | |
| Hexachlorobutadiene | ND | 10 | | | | | | | | |
| Hexachlorocyclopentadiene | ND | 10 | | | | | | | | |
| Hexachloroethane | ND | 10 | | | | | | | | |
| Indeno(1,2,3-cd)pyrene | ND | 10 | | | | | | | | |
| Isophorone | ND | 10 | | | | | | | | |
| 1-Methylnaphthalene | ND | 10 | | | | | | | | |
| 2-Methylnaphthalene | ND | 10 | | | | | | | | |
| 2-Methylphenol | ND | 10 | | | | | | | | |
| 3+4-Methylphenol | ND | 10 | | | | | | | | |
| N-Nitrosodi-n-propylamine | ND | 10 | | | | | | | | |
| N-Nitrosodimethylamine | ND | 10 | | | | | | | | |
| N-Nitrosodiphenylamine | ND | 10 | | | | | | | | |
| Naphthalene | ND | 10 | | | | | | | | |
| 2-Nitroaniline | ND | 10 | | | | | | | | |
| 3-Nitroaniline | ND | 10 | | | | | | | | |
| 4-Nitroaniline | ND | 10 | | | | | | | | |
| Nitrobenzene | ND | 10 | | | | | | | | |
| 2-Nitrophenol | ND | 10 | | | | | | | | |
| 4-Nitrophenol | ND | 10 | | | | | | | | |
| Pentachlorophenol | ND | 20 | | | | | | | | |
| Phenanthrene | ND | 10 | | | | | | | | |
| Phenol | ND | 10 | | | | | | | | |
| Pyrene | ND | 10 | | | | | | | | |
| Pyridine | ND | 10 | | | | | | | | |
| 1,2,4-Trichlorobenzene | ND | 10 | | | | | | | | |
| 2,4,5-Trichlorophenol | ND | 10 | | | | | | | | |
| 2,4,6-Trichlorophenol | ND | 10 | | | | | | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology

Project: KAFB BFF

| Sample ID: mb-29601 | SampType: MBLK | TestCode: EPA Method 8270C: Semivolatiles | | | | | | | | |
|----------------------------|---------------------------------|--|--------------------|-------------|------|----------|-----------|------|----------|------|
| Client ID: PBW | Batch ID: 29601 | RunNo: 40049 | | | | | | | | |
| Prep Date: 1/9/2017 | Analysis Date: 1/13/2017 | SeqNo: 1255089 | Units: µg/L | | | | | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Surr: 2-Fluorophenol | 97 | | 200.0 | | 48.4 | 15 | 98.1 | | | |
| Surr: Phenol-d5 | 77 | | 200.0 | | 38.4 | 15 | 80.7 | | | |
| Surr: 2,4,6-Tribromophenol | 140 | | 200.0 | | 70.4 | 15 | 112 | | | |
| Surr: Nitrobenzene-d5 | 71 | | 100.0 | | 70.7 | 27.2 | 90.7 | | | |
| Surr: 2-Fluorobiphenyl | 64 | | 100.0 | | 63.6 | 23.3 | 85.6 | | | |
| Surr: 4-Terphenyl-d14 | 68 | | 100.0 | | 68.2 | 27.6 | 107 | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology
Project: KAFB BFF

| | | | | | | | | | | |
|------------|-----------------|----------------|------------------|-------------|---------------------------------|----------|-------------|------|----------|------|
| Sample ID | MB-29607 | SampType: | MBLK | TestCode: | EPA Method 7470: Mercury | | | | | |
| Client ID: | PBW | Batch ID: | 29607 | RunNo: | 39928 | | | | | |
| Prep Date: | 1/9/2017 | Analysis Date: | 1/10/2017 | SeqNo: | 1251289 | Units: | mg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Mercury | ND | 0.00020 | | | | | | | | |

| | | | | | | | | | | |
|------------|------------------|----------------|------------------|-------------|---------------------------------|----------|-------------|------|----------|------|
| Sample ID | LCS-29607 | SampType: | LCS | TestCode: | EPA Method 7470: Mercury | | | | | |
| Client ID: | LCSW | Batch ID: | 29607 | RunNo: | 39928 | | | | | |
| Prep Date: | 1/9/2017 | Analysis Date: | 1/10/2017 | SeqNo: | 1251290 | Units: | mg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Mercury | 0.0052 | 0.00020 | 0.005000 | 0 | 105 | 80 | 120 | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

QC SUMMARY REPORT

Hall Environmental Analysis Laboratory, Inc.

WO#: 1701260

17-Jan-17

Client: EA Engineering Science & Technology

Project: KAFB BFF

| Sample ID | MB-29596 | SampType: | MBLK | TestCode: | EPA 6010B: Total Recoverable Metals | | | | | |
|------------|------------------|----------------|------------------|-------------|--|----------|-------------|------|----------|------|
| Client ID: | PBW | Batch ID: | 29596 | RunNo: | 39943 | | | | | |
| Prep Date: | 1/10/2017 | Analysis Date: | 1/10/2017 | SeqNo: | 1251576 | Units: | mg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Arsenic | ND | 0.020 | | | | | | | | |
| Barium | ND | 0.020 | | | | | | | | |
| Cadmium | ND | 0.0020 | | | | | | | | |
| Chromium | ND | 0.0060 | | | | | | | | |
| Lead | ND | 0.0050 | | | | | | | | |
| Selenium | ND | 0.050 | | | | | | | | |
| Silver | ND | 0.0050 | | | | | | | | |

| Sample ID | LCS-29596 | SampType: | LCS | TestCode: | EPA 6010B: Total Recoverable Metals | | | | | |
|------------|------------------|----------------|------------------|-------------|--|----------|-------------|------|----------|------|
| Client ID: | LCSW | Batch ID: | 29596 | RunNo: | 39943 | | | | | |
| Prep Date: | 1/10/2017 | Analysis Date: | 1/10/2017 | SeqNo: | 1251577 | Units: | mg/L | | | |
| Analyte | Result | PQL | SPK value | SPK Ref Val | %REC | LowLimit | HighLimit | %RPD | RPDLimit | Qual |
| Arsenic | 0.45 | 0.020 | 0.5000 | 0 | 90.5 | 80 | 120 | | | |
| Barium | 0.46 | 0.020 | 0.5000 | 0 | 91.0 | 80 | 120 | | | |
| Cadmium | 0.45 | 0.0020 | 0.5000 | 0 | 90.0 | 80 | 120 | | | |
| Chromium | 0.45 | 0.0060 | 0.5000 | 0 | 90.7 | 80 | 120 | | | |
| Lead | 0.44 | 0.0050 | 0.5000 | 0 | 87.6 | 80 | 120 | | | |
| Selenium | 0.44 | 0.050 | 0.5000 | 0 | 87.9 | 80 | 120 | | | |
| Silver | 0.092 | 0.0050 | 0.1000 | 0 | 91.7 | 80 | 120 | | | |

Qualifiers:

- * Value exceeds Maximum Contaminant Level.
- D Sample Diluted Due to Matrix
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- R RPD outside accepted recovery limits
- S % Recovery outside of range due to dilution or matrix
- B Analyte detected in the associated Method Blank
- E Value above quantitation range
- J Analyte detected below quantitation limits
- P Sample pH Not In Range
- RL Reporting Detection Limit
- W Sample container temperature is out of limit as specified

Sample Log-In Check List

Client Name: EA Engineering Alb

Work Order Number: 1701260

RcptNo: 1

Received by/date: AT 01/09/17

Logged By: Anne Thorne 1/9/2017 9:25:00 AM *Anne Thorne*

Completed By: Anne Thorne 1/9/2017 11:10:03 AM *Anne Thorne*

Reviewed By: *AT 01/09/17*

Chain of Custody

- 1. Custody seals intact on sample bottles? Yes No Not Present
- 2. Is Chain of Custody complete? Yes No Not Present
- 3. How was the sample delivered? Client

Log In

- 4. Was an attempt made to cool the samples? Yes No NA
- 5. Were all samples received at a temperature of >0° C to 6.0°C Yes No NA
- 6. Sample(s) in proper container(s)? Yes No
- 7. Sufficient sample volume for indicated test(s)? Yes No
- 8. Are samples (except VOA and ONG) properly preserved? Yes No
- 9. Was preservative added to bottles? Yes No NA
- 10. VOA vials have zero headspace? Yes No No VOA Vials
- 11. Were any sample containers received broken? Yes No
- 12. Does paperwork match bottle labels? Yes No
 (Note discrepancies on chain of custody)
- 13. Are matrices correctly identified on Chain of Custody? Yes No
- 14. Is it clear what analyses were requested? Yes No
- 15. Were all holding times able to be met? Yes No
 (If no, notify customer for authorization.)

of preserved bottles checked for pH: 1, 2
 (<2 or >12 unless noted)
 Adjusted? NO
 Checked by: AT

Special Handling (if applicable)

- 16. Was client notified of all discrepancies with this order? Yes No NA

Person Notified: _____ Date: _____
 By Whom: _____ Via: eMail Phone Fax In Person
 Regarding: _____
 Client Instructions: _____


17. Additional remarks:

18. Cooler Information

| Cooler No | Temp °C | Condition | Seal Intact | Seal No | Seal Date | Signed By |
|-----------|---------|-----------|-------------|---------|-----------|-----------|
| 1 | 1.0 | Good | Yes | | | |

Chain-of-Custody Record

Turn-Around Time: _____
 Standard Rush
 Project Name: **KAFB - BFF**
 Project #: **62599DM01.1028.11**
 Project Manager: **Devon Jercinovic**
 Sampler: **Joshua Brown**
 On Ice: Yes No
 Sample Temperature: **1.00C**



HALL ENVIRONMENTAL ANALYSIS LABORATORY

www.hallenvironmental.com

4901 Hawkins NE - Albuquerque, NM 87109

Tel. 505-345-3975 Fax 505-345-4107

Client: **EA Engineering Engineering**
 Mailing Address: _____
 Phone #: _____
 email or Fax#: _____
 QA/QC Package:
 Standard Level 4 (Full Validation)
 Accreditation
 NELAP Other _____
 EDD (Type) _____

Analysis Request

| Date | Time | Matrix | Sample Request ID | Container Type and # | Preservative Type | HEAL No. | BTEX + MTBE + TMB's (8021) | BTEX + MTBE + TPH (Gas only) | TPH 8015B (GRO / DRO / MRO) | TPH (Method 418.1) | EDB (Method 504.1) | PAH's (8310 or 8270 SIMS) | RCRA 8 Metals + Hg | Anions (F, Cl, NO ₃ , NO ₂ , PO ₄ , SO ₄) | 8081 Pesticides / 8082 PCB's | 8260B (VOA) | 8270 (Semi-VOA) | RCI | Air Bubbles (Y or N) |
|--------|-------|--------|-------------------|----------------------|-------------------|----------|----------------------------|------------------------------|-----------------------------|--------------------|--------------------|---------------------------|--------------------|--|------------------------------|-------------|-----------------|-----|----------------------|
| 1-8-16 | 12:10 | water | KAFB-106239 LC | | | 1701260 | | | | | | | X | | | X | X | X | |
| | | | Trip Blank | | | -002 | | | | | | | | | X | | | | |
| | | | KC 01109117 | | | | | | | | | | | | | | | | |

Date: **1-9-17** Time: **0925** Relinquished by: *[Signature]*
 Received by: *[Signature]* Date: **1/9/17** Time: **0925**
 Date: _____ Time: _____ Relinquished by: _____
 Received by: _____ Date: _____ Time: _____

Remarks: **Email analytical to jbrown@east.com + emorse@east.com**
Thanks

If necessary, samples submitted to Hall Environmental may be subcontracted to other accredited laboratories. This serves as notice of this possibility. Any sub-contracted data will be clearly notated on the analytical report.





