


***** Retransmission *****



2709-D Pan American Freeway NE
Albuquerque, New Mexico 87107
Phone (505) 344-3777
Fax (505) 344-4413

FAX TRANSMITTAL SHEET

DELIVER TO: Dave

PHONE NUMBER: _____

COMPANY: 827 1544

FAX NUMBER: _____

NUMBER OF PAGES BEING SENT: 4 (INCLUDING THIS PAGE)

From:

Date: 6/17

- H. Mitchell Rubenstein, Ph.D., President/CEO
- Van K. Barta, CBC
- Kimberly D. McNeill, Project Manager
- Jacinta A. Tenorio, Administrative Assistant
- Francine J. Torivio, Sample Control

Time: 4:30

FAX NUMBER:
(505) 344-4413

This information contained in this Facsimile message is intended only for the personal and confidential use of the designated recipients named above. This message may be proprietary confidential information of a client, and as such is privileged and confidential. If the reader of this message is not the intended recipient or any agent responsible for delivering it to the intended recipient, you are hereby notified that you have received this document in error, and that any review, dissemination, distribution or copying of this message is strictly prohibited. If you have received this communication in error, please notify us immediately by telephone and return the original message to us by mail. Thank you.....

COMMENTS:

8021 Results

If you did not receive all pages of this transmission or if you experience Fax Transmission problems. Please call (505) 344-3777, as soon as possible after the receipt.

GAS CHROMATOGRAPHY RESULTS

TEST : 8021 HALO
 CLIENT : NMED-HRMB
 PROJECT # : (none)
 PROJECT NAME : NMED-PNM

PINNACLE I.D.: 906016

ID. #	CLIENT I.D.	MATRIX	DATE SAMPLED	DATE EXTRACTED	DATE ANALYZED	DIL. FACTOR
07	PSMW-D	AQUEOUS	6/2/99	NA	6/7/99	1
08	TRIP BLANK	AQUEOUS	5/22/99	NA	6/7/99	1
PARAMETER	DET. LIMIT	UNITS	PSMW-D	TRIP BLANK		
BROMODICHLOROMETHANE	0.2	UG/L	< 0.2	< 0.2		
BROMOFORM	0.5	UG/L	< 0.5	< 0.5		
BROMOMETHANE	1.0	UG/L	< 1.0	< 1.0		
CARBON TETRACHLORIDE	0.2	UG/L	< 0.2	< 0.2		
CHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5		
CHLOROETHANE	0.5	UG/L	< 0.5	< 0.5		
CHLOROFORM	0.5	UG/L	< 0.5	< 0.5		
CHLOROMETHANE	1.0	UG/L	2.0	< 0.5		
DIBROMOCHLOROMETHANE	0.2	UG/L	< 1.0	< 1.0		
1,2-DIBROMOETHANE (EDB)	0.2	UG/L	< 0.2	< 0.2		
1,2-DICHLOROBENZENE	0.5	UG/L	< 0.2	< 0.2		
1,3-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5		
1,4-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5		
1,1-DICHLOROETHANE	0.3	UG/L	< 0.5	< 0.5		
1,2-DICHLOROETHANE (EDC)	0.5	UG/L	3.8	< 0.3		
1,1-DICHLOROETHENE	0.2	UG/L	< 0.5	< 0.5		
cis-1,2-DICHLOROETHENE	0.2	UG/L	42	< 0.2		
trans-1,2-DICHLOROETHENE	1.0	UG/L	0.2	< 0.2		
1,2-DICHLOROPROPANE	0.2	UG/L	< 1.0	< 1.0		
cis-1,3-DICHLOROPROPENE	0.2	UG/L	< 0.2	< 0.2		
trans-1,3-DICHLOROPROPENE	0.2	UG/L	< 0.2	< 0.2		
METHYLENE CHLORIDE	2.0	UG/L	< 0.2	< 0.2		
1,1,2-TETRACHLOROETHANE	0.5	UG/L	< 2.0	< 2.0		
TETRACHLOROETHENE	0.5	UG/L	< 0.5	< 0.5		
1,1,1-TRICHLOROETHANE	1.0	UG/L	160	< 0.5		
1,1,2-TRICHLOROETHANE	0.2	UG/L	6.9	< 1.0		
TRICHLOROETHENE	0.3	UG/L	< 0.2	< 0.2		
TRICHLOROFLUOROMETHANE	0.2	UG/L	0.5	< 0.3		
VINYL CHLORIDE	0.5	UG/L	< 0.2	< 0.2		
			< 0.5	< 0.5		
SURROGATE:						
BROMOCHLOROMETHANE (%)			108	107		
SURROGATE LIMITS	(73 - 117)					
TRIFLUOROTOLUENE (%)			116	112		
SURROGATE LIMITS	(69 - 117)					

CHEMIST NOTES:

N/A

GAS CHROMATOGRAPHY RESULTS

TEST : 8021 HALO
 CLIENT : NMED-HRMB
 PROJECT # : (none)
 PROJECT NAME : NMED-PNM

PINNACLE I.D.: 906016

SAMPLE						
ID. #	CLIENT I.D.	MATRIX	DATE SAMPLED	DATE EXTRACTED	DATE ANALYZED	DIL. FACTOR
01	PSMW-01R	AQUEOUS	6/2/00	NA	6/7/99	1
02	PSMW-8A	AQUEOUS	6/2/99	NA	6/7/99	1
03	PSMW-10	AQUEOUS	6/2/99	NA	6/7/99	1
PARAMETER	DET. LIMIT	UNITS	PSMW-01R	PSMW-8A	PSMW-10	
BROMODICHLOROMETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
BROMOFORM	0.5	UG/L	< 0.5	< 0.5	< 0.5	
BROMOMETHANE	1.0	UG/L	< 1.0	< 1.0	< 1.0	
CARBON TETRACHLORIDE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
CHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5	
CHLOROETHANE	0.5	UG/L	< 0.5	< 0.5	< 0.5	
CHLOROFORM	0.5	UG/L	1.3	1.0	1.4	
CHLOROMETHANE	1.0	UG/L	< 1.0	< 1.0	< 1.0	
DIBROMOCHLOROMETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
1,2-DIBROMOETHANE (EDB)	0.2	UG/L	< 0.2	< 0.2	< 0.2	
1,2-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5	
1,3-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5	
1,4-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5	
1,1-DICHLOROETHANE	0.3	UG/L	3.3	2.7	2.9	
1,2-DICHLOROETHANE (EDC)	0.5	UG/L	< 0.5	< 0.5	< 0.5	
1,1-DICHLOROETHENE	0.2	UG/L	14	19	36	
cis-1,2-DICHLOROETHENE	0.2	UG/L	0.2	0.4	0.2	
trans-1,2-DICHLOROETHENE	1.0	UG/L	< 1.0	< 1.0	< 1.0	
1,2-DICHLOROPROPANE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
cis-1,3-DICHLOROPROPENE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
trans-1,3-DICHLOROPROPENE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
METHYLENE CHLORIDE	2.0	UG/L	< 2.0	< 2.0	< 2.0	
1,1,2,2-TETRACHLOROETHANE	0.5	UG/L	< 0.5	< 0.5	< 0.5	
TETRACHLOROETHENE	0.5	UG/L	42	40	140	
1,1,1-TRICHLOROETHANE	1.0	UG/L	3.5	4.0	5.4	
1,1,2-TRICHLOROETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
TRICHLOROETHENE	0.3	UG/L	< 0.3	< 0.3	0.4	
TRICHLOROFLUOROMETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2	
VINYL CHLORIDE	0.5	UG/L	< 0.5	< 0.5	< 0.5	

SURROGATE:

BROMOCHLOROMETHANE (%)		106	108	94
SURROGATE LIMITS	(73 - 117)			
TRIFLUOROTOLUENE (%)		113	107	99
SURROGATE LIMITS	(69 - 117)			

CHEMIST NOTES:

N/A

GAS CHROMATOGRAPHY RESULTS

TEST : 8021 HALO
 CLIENT : NMED-HRMB
 PROJECT # : (none)
 PROJECT NAME : NMED-PNM

PINNACLE I.D.: 900010

SAMPLE		DATE	DATE	DATE	DIL.
ID. #	CLIENT I.D.	SAMPLED	EXTRACTED	ANALYZED	FACTOR
04	PSMW-16	6/2/99	NA	6/7/99	1
05	PSMW-27	6/2/99	NA	6/7/99	1
06	PSMW-27-500	6/2/99	NA	6/7/99	1

PARAMETER	DET. LIMIT	UNITS	PSMW-16	PSMW-27	PSMW-27-500
BROMODICHLOROMETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2
BROMOFORM	0.5	UG/L	< 0.5	< 0.5	< 0.5
BROMOMETHANE	1.0	UG/L	< 1.0	< 1.0	< 1.0
CARBON TETRACHLORIDE	0.2	UG/L	< 0.2	< 0.2	< 0.2
CHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5
CHLOROETHANE	0.5	UG/L	< 0.5	< 0.5	< 0.5
CHLOROFORM	0.5	UG/L	1.1	1.2	5.5
CHLOROMETHANE	1.0	UG/L	< 1.0	< 1.0	< 1.0
DIBROMOCHLOROMETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2
1,2-DIBROMOETHANE (EDB)	0.2	UG/L	< 0.2	< 0.2	< 0.2
1,2-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5
1,3-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5
1,4-DICHLOROBENZENE	0.5	UG/L	< 0.5	< 0.5	< 0.5
1,1-DICHLOROETHANE	0.3	UG/L	1.3	< 0.3	< 0.3
1,2-DICHLOROETHANE (EDC)	0.5	UG/L	< 0.5	< 0.5	< 0.5
1,1-DICHLOROETHENE	0.2	UG/L	25	16	5.4
cis-1,2-DICHLOROETHENE	0.2	UG/L	< 0.2	< 0.2	< 0.2
trans-1,2-DICHLOROETHENE	1.0	UG/L	< 1.0	< 1.0	< 1.0
1,2-DICHLOROPROPANE	0.2	UG/L	< 0.2	< 0.2	< 0.2
cis-1,3-DICHLOROPROPENE	0.2	UG/L	< 0.2	< 0.2	< 0.2
trans-1,3-DICHLOROPROPENE	0.2	UG/L	< 0.2	< 0.2	< 0.2
METHYLENE CHLORIDE	2.0	UG/L	< 2.0	< 2.0	< 2.0
1,1,2,2-TETRACHLOROETHANE	0.5	UG/L	< 0.5	< 0.5	< 0.5
TETRACHLOROETHENE	0.5	UG/L	78	13	2.1
1,1,1-TRICHLOROETHANE	1.0	UG/L	1.9	< 1.0	< 1.0
1,1,2-TRICHLOROETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2
TRICHLOROETHENE	0.3	UG/L	< 0.3	< 0.3	< 0.3
TRICHLOROFLUOROMETHANE	0.2	UG/L	< 0.2	< 0.2	< 0.2
VINYL CHLORIDE	0.5	UG/L	< 0.5	< 0.5	< 0.5

SURROGATE:

BROMOCHLOROMETHANE (%)		104	104	102
SURROGATE LIMITS	(73 - 117)			
TRIFLUOROTOLUENE (%)		106	NA	NA
SURROGATE LIMITS	(69 - 117)			

CHEMIST NOTES:

N/A