



2709-D Pan American Freeway NE
 Albuquerque, New Mexico 87107
 Phone (505) 344-3777
 Fax (505) 344-4413

Pinnacle Lab ID number **906016**
 July 16, 1999

NMED-HRMB
 2044A GALISTEO ST.
 SANTA FE, NM 87505

Project Name NMED-PNM
 Project Number (none)

Attention: CARL WILL

On 6/2/99 Pinnacle Laboratories, Inc. Inc., (ADHS License No. AZ0592), received a request to analyze **aqueous** samples. The samples were analyzed with EPA methodology or equivalent methods. The results of these analyses and the quality control data, which follow each set of analyses, are enclosed.

This report is being reissued in part to correct an error on the Lead result. We apologize for any inconvenience this may have caused. This report was originally dated 6/24/99.

EPA method 8021 was performed by Pinnacle Laboratories, Inc., Albuquerque, NM.

All other parameters were performed by ESL (OR) Inc., Portland, OR.

If you have any questions or comments, please do not hesitate to contact us at (505)344-3777.

Kimberly D. McNeill
 Project Manager

H. Mitchell Rubenstein, Ph. D.
 General Manager

MR: mt

Enclosure

Environmental Services Laboratory, Inc.



17400 SW Upper Boones Ferry Road • Suite 270 • Portland, OR 97224 • (503) 670-8520

July 02, 1999

Kim McNeill
Pinnacle Laboratories
2709-D Pan American Fwy NE
Albuquerque, NM 87107

TEL: 505-344-3777
FAX (505) 344-4413

RE: 906016/NMHR/NMED-PNM

Order No.: 9906022

Dear Kim McNeill,

Environmental Services Laboratory received 1 sample on 06/03/99 for the analyses presented in the following report.

The Samples were analyzed for the following tests:

Hexavalent Chrome (SW7196A)
ICP Metals (ICPMET)

There were no problems with the analyses and all data for associated QC met EPA or laboratory specifications except where noted in the Case Narrative. Results apply only to the samples analyzed. Reproduction of this report is permitted only in its entirety, without the written approval from the Laboratory.

If you have any questions regarding these tests results, please feel free to call.

Sincerely,

Kimberly Hill
Project Manager

Technical Review

Environmental Services Laboratory

Date: 02-Jul-99

CLIENT: Pinnacle Laboratories

Client Sample ID: 906016-02

Lab Order: 9906022

Tag Number:

Project: 906016/NMHR/NMED-PNM

Collection Date: 06/02/99

Lab ID: 9906022-01A

Matrix: AQUEOUS

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
HEXAVALENT CHROME		8023				Analyst: sld
Hexavalent Chrome	0.081	0.005		mg/L	1	06/03/99
ICP METALS		SW 6010 / EPA 200.				Analyst: btn
Chromium	0.081	0.05		mg/L	10	06/17/99
Lead	ND	0.005		mg/L	10	06/17/99

Qualifiers:
ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level

S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range