


Ciniza



COVER LETTER

Tuesday, July 25, 2006

Steve Morris
Giant Refining Co
Rt. 3 Box 7
Gallup, NM 87301

TEL: (505) 722-3833
FAX (505) 722-0210

RE: Boundary Well 2B for Selenium

Order No.: 0607244

Dear Steve Morris:

Hall Environmental Analysis Laboratory, Inc. received 1 sample(s) on 7/21/2006 for the analyses presented in the following report.

These were analyzed according to EPA procedures or equivalent.

Reporting limits are determined by EPA methodology. No determination of compounds below these (denoted by the ND or < sign) has been made.

Please don't hesitate to contact HEAL for any additional information or clarifications.


Sincerely,



Andy Freeman, Business Manager
Nancy McDuffie, Laboratory Manager

AZ license # AZ0682
ORELAP Lab # NM100001





Hall Environmental Analysis Laboratory, Inc.

Date: 25-Jul-06

CLIENT: Giant Mining Co Client Sample ID: BW-2B
 Lab Order: 0607244 Collection Date: 7/19/2006 3:30:00 PM
 Project: Boundary Well 2B for Selenium Date Received: 7/21/2006
 Lab ID: 0607244-01 Matrix: AQUEOUS

Analyses	Result	PQL	Qual	Units	DF	Date Analyzed
EPA METHOD 6010: DISSOLVED METALS						
Selenium	ND	0.050		mg/L	1	Analyst: NMO 7/25/2006 11:36:17 AM
EPA 6010: TOTAL RECOVERABLE METALS						
Selenium	ND	0.050		mg/L	1	Analyst: NMO 7/25/2006 9:13:00 AM

- Qualifiers:
- * Value exceeds Maximum Contaminant Level
 - E Value above quantitation range
 - J Analyte detected below quantitation limits
 - S Spike Recovery outside accepted recovery limits
 - B Analyte detected in the associated Method Blank
 - H Holding times for preparation or analysis exceeded
 - ND Not Detected at the Reporting Limit

QA/QC SUMMARY REPORT

Client: Giant Refining Co
 Project: Boundary Well 2B for Selenium

Work Order: 0607244

Analyte	Result	Units	PQL	%Rec	LowLimit	HighLimit	%RPD	RPDLimit	Qual
---------	--------	-------	-----	------	----------	-----------	------	----------	------

Method: SW6010A									
Sample ID: MB		MBLK							
Selenium	ND	mg/L	0.050						
Sample ID: LCS		LCS							
Selenium	0.4790	mg/L	0.050	95.8	80	120			

Method: SW6010A									
Sample ID: 0607244-01AMSD		MSD							
Selenium	0.4837	mg/L	0.050	96.7	75	125	0.983	20	
Sample ID: MB-10853		MBLK							
Selenium	ND	mg/L	0.050						
Sample ID: LCS-10853		LCS							
Selenium	0.4717	mg/L	0.050	94.3	80	120			
Sample ID: 0607244-01AMS		MS							
Selenium	0.4790	mg/L	0.050	95.8	75	125			

Qualifiers:

- E Value above quantitation range
- J Analyte detected below quantitation limits
- R RPD outside accepted recovery limits
- H Holding times for preparation or analysis exceeded
- ND Not Detected at the Reporting Limit
- S Spike Recovery outside accepted recovery limits

Hall Environmental Analysis Laboratory, Inc.

Sample Receipt Checklist

Client Name GIANTREFIN

Date and Time Received:

7/21/2006

Work Order Number 0607244

Received by AT

Checklist completed by

[Handwritten Signature]

7/21/06

Signature

Date

Matrix

Carrier name FedEx

- Shipping container/cooler in good condition? Yes No Not Present
- Custody seals intact on shipping container/cooler? Yes No Not Present Not Shipped
- Custody seals intact on sample bottles? Yes No N/A
- Chain of custody present? Yes No
- Chain of custody signed when relinquished and received? Yes No
- Chain of custody agrees with sample labels? Yes No
- Samples in proper container/bottle? Yes No
- Sample containers intact? Yes No
- Sufficient sample volume for indicated test? Yes No
- All samples received within holding time? Yes No
- Water - VOA vials have zero headspace? No VOA vials submitted Yes No
- Water - pH acceptable upon receipt? Yes No N/A

Container/Temp Blank temperature? 5° 4° C ± 2 Acceptable
If given sufficient time to cool.

COMMENTS:

Client contacted _____ Date contacted: _____ Person contacted _____

Contacted by: _____ Regarding _____

Comments: _____

Corrective Action _____

