



Laboratory Report Number: L12010479

Mark Lyon
Environmental Waste Solutions
2440 Louisiana Blvd
Albuquerque, NM 87110

Please find enclosed the analytical results for the samples you submitted to Microbac Laboratories. Review and compilation of your report was completed by Microbac's Ohio Valley Division (OVD). If you have any questions, comments, or require further assistance regarding this report, please contact your service representative listed below.

Laboratory Contact:
Stephanie Mossburg – Team Chemist/Data Specialist
(740) 373-4071
Stephanie.Mossburg@microbac.com

I certify that all test results meet all of the requirements of the DoD QSM and other applicable contract terms and conditions. Any exceptions are attached to this cover page or addressed in the method narratives presented in the report. All results for soil samples are reported on a 'dry-weight' basis unless specified otherwise. Analytical results for water and wastes are reported on a 'as received' basis unless specified otherwise. A statement of uncertainty for each analysis is available upon request. This laboratory report shall not be reproduced, except in full, without the written approval of Microbac Laboratories, DoD ELAP certification number 2936.01. The reported results are related only to the samples analyzed as received.

This report was certified on February 02 2012

David Vandenberg – Managing Director

State of Origin: NM
Accrediting Authority: N/A ID:N/A
QAPP: DOD Ver 4.1



Record of Sample Receipt and Inspection

Comments/Discrepancies

This is the record of the shipment conditions and the inspection records for the samples received and reported as a sample delivery group (SDG). All of the samples were inspected and observed to conform to our receipt policies, except as noted below.

There were no discrepancies.

Discrepancy	Resolution

Coolers

Cooler #	Temperature Gun	Temperature	COC #	Airbill #
0016941	H	3.0		1002239502710004575000874824308790

Inspection Checklist

#	Question	Result
1	Were shipping coolers sealed?	Yes
2	Were custody seals intact?	Yes
3	Were cooler temperatures in range of 0-6?	Yes
4	Was ice present?	Yes
5	Were COC's received/information complete/signed and dated?	Yes
6	Were sample containers intact and match COC?	Yes
7	Were sample labels intact and match COC?	Yes
8	Were the correct containers and volumes received?	Yes
9	Were samples received within EPA hold times?	Yes
10	Were correct preservatives used? (water only)	Yes
11	Were pH ranges acceptable? (voa's excluded)	Yes
12	Were VOA samples free of headspace (less than 6mm)?	NA

Lab Report #: L12010479

Lab Project #: 3005.011

Project Name: White Sands MR

Lab Contact: Stephanie Mossburg

Samples Received

Client ID	Laboratory ID	Date Collected	Date Received
MP05-0112-1	L12010479-01	01/18/2012 15:15	01/19/2012 10:43



Login Number: L12010479
Department: Conventionals
Analyst: Dorothy Payne

METHOD

Analysis SW846 9040C,9045D/EPA 150.1/SM4500-H B (pH)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41460

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Metals
Analyst: Kim Rhodes

METHOD

Preparation: SW-846 3005

Analysis: SW-846 6010

HOLDING TIMES

Sample Preparation: All holding times were met.

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

CALIBRATION

Initial Calibration: All acceptance criteria were met.

Alternate Source Standards: All acceptance criteria were met.

Interference Check Standards: All acceptance criteria were met.

Continuing Calibration Verification: All acceptance criteria were met.

Continuing Calibration Blank: All acceptance criteria were met.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Serial Dilution/Post Digestion Spikes: WG387580 - All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: WG387580 - Calcium for client sample 01 was reanalyzed at a dilution in order to obtain a result within the calibration range.

Narrative ID: 41260

Approved By: Maren Beery

Maren Beery



Login Number: L12010479
Department: Metals
Analyst: Erin Long

METHOD

Preparation: SW-846 3015

Analysis: SW-846 6020

HOLDING TIMES

Sample Preparation: All holding times were met.

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

CALIBRATION

Initial Calibration: All acceptance criteria were met.

Alternate Source Standards: All acceptance criteria were met.

Interference Check Standards: All acceptance criteria were met.

Continuing Calibration: All acceptance criteria were met.

Continuing Calibration Blank: WG387531 - Due to continuing calibration blank failure for thallium on 20-JAN-2012 at 10:39, client sample 01 and all batch QA/QC were reanalyzed on a later calibration which was compliant for thallium.

Low Level Check: WG387531 - Due to low level initial calibration verification failure for cadmium on 20-JAN-2012 at 09:53, client sample 01 and all batch QA/QC were reanalyzed on a later calibration which was compliant for cadmium.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Serial Dilution/Post Digestion Spikes: WG387531 - All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: WG387531 - Due to suspected high levels of nontarget analytes, client sample 01 was analyzed initially at a dilution for all analytes. However, the sample yielded non-detect results in the initial analysis. Therefore, client sample 01 was reanalyzed undiluted for all analytes.

Narrative ID: 41140

Approved By: Maren Beery

Maren Beery



Login Number: L12010479
Department: Metals - AA
Analyst: Pierce Morris

METHOD

Preparation: SW-846 7470

Analysis: SW-846 7470

HOLDING TIMES

Sample Preparation: All holding times were met.

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

CALIBRATION

Initial Calibration: All acceptance criteria were met.

Alternate Source Standards: All acceptance criteria were met.

Interference Check Standards: All acceptance criteria were met.

Continuing Calibration Verification: All acceptance criteria were met.

Continuing Calibration Blank: All acceptance criteria were met.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Serial Dilution/Post Digestion Spikes: WG387814 - All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

Narrative ID: 41527

Approved By: Maren Beery

Maren Beery



Login Number: L12010479
Department: General Chromatography
Analyst: Jeremy Kinney

METHOD

Analysis SW-846 9056/300.0

HOLDING TIMES

Sample Preparation: All holding times were met.

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

CALIBRATION

Initial Calibration: All acceptance criteria were met.

Alternate Source Standards: All acceptance criteria were met.

Continuing Calibration and Tune: All acceptance criteria were met.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

Surrogates: All acceptance criteria were met.

Manual Integration Reason Codes

Reason #1: Data System Fails to Select Correct Peak In some cases the chromatography system selects and integrates the 'wrong peak'. In this case the analyst must correct the selection and force the system to integrate the proper peak. Other times the system may miss the peak completely.

Reason #2: Data System Splits the Peak Incorrectly or Integrates a False Peak as a Rider Peak This phenomena is common at low concentrations where the signal:noise ratio is low. A single compound (peak) is incorrectly split into multiple peaks or integrated as a main peak with one or more rider peaks resulting in low area counts for the target compound.

Reason #3: Improperly Integrated Isomers and/or coeluting compounds. This system often fails to distinguish coeluting compounds and or isomers. The integration areas and concentrations are wrong, and they must be corrected by manual integration. Prime examples are benzo(k)fluoranthene and benzo(b)fluoranthene which are often unresolved and integrated improperly when both are present at low concentrations in standards or samples.

Reason #4: System Establishes Incorrect Baseline There are numerous situations in chromatography where the system establishes the baseline incorrectly. Some baseline errors will be obvious to the analyst and should be corrected via manual procedures.

Reason #5: Miscellaneous Other situations involving integration errors may require in-depth review and technical judgment. These cases should be brought to the attention of the laboratory management. If the form of manual integration is not clearly covered by these four cases, then review and approval by the Laboratory Director or the QA/QC Supervisor will be required.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41294

Approved By: Mike Cochran





Login Number: L12010479
Department: Conventionals
Analyst: Deanna Hesson

METHOD

Analysis EPA 310.2 (Alkalinity)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: The blank result was lower than the LOQ (reporting limit), but greater than the LOD (MDL).

Laboratory Control Sample: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41109

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Jeremy Kinney

METHOD

Analysis SW846 9014/9010C/SM4500-CN-C,E-20th (Cyanide)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: Cyanide-Amenable is the difference between the total cyanide and the treated cyanide. The LCS is analyzed to show that all of the cyanide is amenable (the treated portion is ND). The LCS forms cannot calculate cyanide amenable. The LCS is acceptable.

Matrix Spikes: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41402

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Dorothy Payne

METHOD

Analysis EPA 120.1/SM2510 B (Conductivity)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41459

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Deanna Hesson

METHOD

Analysis EPA 350.1/SM4500-NH3 B(NH3)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41461

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Deanna Hesson

METHOD

Analysis EPA 353.2/SM4500-NO3 F (Nitrate)

HOLDING TIMES

Sample Analysis: Nitrate is reported as the difference of nitrate-nitrite (28 day hold) and nitrite (48 hour hold). Both analysis were analyzed within the appropriate hold time. The nitrate hold time is within compliance.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41462

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Dorothy Payne

METHOD

Analysis EPA 365.2/SM4500-P E (Orthophosphate)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41463

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Holly Reed

METHOD

Analysis EPA 160.1/SM2540 C(Total Dissolved Solids)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41465

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Deanna Hesson

METHOD

Analysis Water: EPA 415.1/SM5310C/SW846 9060 (Total Organic Carbon)
Soil: Lloyd-Khan Methodology

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QAI/QC

Method Blank: The blank result was lower than the LOQ (reporting limit), but greater than the LOD (MDL).

Laboratory Control Sample: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41224

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".



Login Number: L12010479
Department: Conventionals
Analyst: Holly Reed

METHOD

Analysis EPA 160.2/SM2540 D (Total Suspended Solids)

HOLDING TIMES

Sample Analysis: All holding times were met.

PREPARATION

Sample preparation proceeded normally.

BATCH QA/QC

Method Blank: All acceptance criteria were met.

Laboratory Control Sample: All acceptance criteria were met.

Duplicates: All acceptance criteria were met.

Matrix Spikes: All acceptance criteria were met.

SAMPLES

Samples: All acceptance criteria were met.

I certify that this data package is in compliance with the terms and conditions agreed to by the client and Microbac Laboratories Inc., both technically and for completeness, except for the conditions noted above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Manager or designated person, as verified by the following signature.

Narrative ID: 41464

Approved By: Deanna Hesson

A handwritten signature in cursive script that reads "Deanna Hesson".

COC No. A 28474

158 Starlite Drive
Marietta, OH 45750



Phone: 740-373-4071
Fax: 740-373-4885

CHAIN-OF-CUSTODY RECORD

Company Name: Zia - Shaw		Contact Phone #: 505-262-8920		Location: WSMR-STP		Signature: Britt Dan		Matrix* Water	
Project Contact: MARK LYON		Turn Around Requirements: Normal		Project ID: STP		Sampler (print): Bradley T. Davis		Sample I.D. No. MP05-OR-1	
Relinquished by: (Signature) Bradley T. Davis	Date 1-18-12	Time 1515	Received (Signal)	NUMBER OF CONTAINERS 7					
Relinquished by: (Signature)	Date	Time	Received (Signal)	Hold					
				Anions					
				Cond pH PO4					
				Alkalinity					
				Total Free Ammonia CN					
				Metals					
				NH3 TOC NO3NO2					
				TSS TDS					
				TOTAL # (LAB USE)					
				Program					
				<input type="checkbox"/> CWA <input type="checkbox"/> RCRA <input type="checkbox"/> DOD <input type="checkbox"/> AFCEE <input type="checkbox"/> Other					
				ADDITIONAL REQUIREMENTS					
				Date					
				Time					
				Received by: (Signature)					
Remarks:			221000021729						

Microbac OVD
 Received: 01/19/2012 10:43
 By: JANE SCHARAD
Jane K. Scharad



Relinquished by: (Signature) **Britt Dan** Date **1-18-12** Time **1700** Received (Signal)

Relinquished by: (Signature) Date Time Received (Signal)

*Water (W), Soil (S), Solid Waste (SD), Unknown (X)