Environmental • Fire • HazMat • CBRNE

FarrWest Environmental Supply

Certificate of Calibration

Calibration Reference Number
101482

Calibration Information

Customer	Date	Technician
ZIA	2013-10-21	John Covington

Device Information

Model	Manufacturer	Serial Number
PRO PLUS	YSI	12c100049

Calibration Standards

Calibration Standard	Mfr.	Exp.	Lot Number
Buffer Solution pH 4	AURICAL	2015-03-30	3032294
Buffer Solution pH 7	AURICAL	2015-01-30	3011582
Buffer Solution pH 10	AURICAL	2014-11-30	2111572
Conductivity 1,413 Âµmhos	AURICAL	2014-05-30	3052442
Redox Buffer (Zobell	AUROCAL	2014-05-30	3043023

Sensor Readings

Sensor:	pH	Cond	Temp	D.O.	ORP	TDS
Reading:	4,7,10	1.41mS/cm	21.5C	8.5mg/L	229mV	

Calibration Result

Passed

Comments

good

The above instrument has been calibrated according to manufacturers's instructions and by comparison to laboratory standards traceable to the National Institute of Standards and Technology (NIST).

FarrWest Environmental Supply
108 Commercial Place, Suite 200 Schertz, Texas 78154
Ph: 210-566-1857 Fax: 210-566-1897 Email: info@farrwestenv.com