



Inorganics Case Narrative

NMED Hazardous Waste Bureau

KAFB - BFF 1Q12

Work Order Number: 1202056

1. This report consists of 3 water samples.
2. The samples were received cool and intact by ALS on 02/07/12.
3. The samples had been correctly preserved for the requested analyses.
4. The samples were prepared for analysis based on Methods for the Chemical Analysis of Waters and Wastes (MCAWW), May 1994 procedures and Environmental Monitoring Systems Laboratory (EMSL) Rev 2.1 procedures.
5. The samples were analyzed following MCAWW and EMSL procedures for the following methods:

<u>Analyte</u>	<u>Method</u>	<u>SOP #</u>
Alkalinity	310.1	1106 Rev 10
Bicarbonate	310.1	1106 Rev 10
Carbonate	310.1	1106 Rev 10
Ammonia as N	350.1	1129 Rev 7
Nitrate/nitrite as N	353.2 Revision 2.0	1127 Rev 8
Sulfide	376.1	1120 Rev 7
Bromide	300.0 Revision 2.1	1113 Rev 12
Chloride	300.0 Revision 2.1	1113 Rev 12
Fluoride	300.0 Revision 2.1	1113 Rev 12
Sulfate	300.0 Revision 2.1	1113 Rev 12

6. All standards and solutions were used within their recommended shelf life.
7. The samples were prepared and analyzed within the established hold time for each analysis.

All in house quality control procedures were followed, as described below.



8. General quality control procedures.

- A preparation (method) blank and laboratory control sample (LCS) were prepared and analyzed with the samples in each preparation batch. There were not more than 20 samples in each preparation batch.
- The method blank associated with each batch was below the reporting limit for the requested analytes. This indicates that no contaminants were introduced to the samples during preparation and analysis.
- The LCS was within the acceptance limits for each analysis.
- All initial and continuing calibration blanks (ICB/CCB) associated with each applicable analytical batch were below the reporting limit for the requested analytes.
- All initial and continuing calibration verifications (ICV/CCV) associated with each applicable analytical batch were within the acceptance criteria for the requested analytes.

9. Matrix specific quality control procedures.

Per method requirements, matrix QC was performed for each analysis. Since a sample from this order number was not the selected quality control (QC) sample, matrix specific QC results are not included in this report.

10. Reduced aliquots were taken of the samples for the alkalinity, bicarbonate, and carbonate analysis. Reporting limits were elevated accordingly.

11. Manual integrations are performed when needed to provide consistent and defensible data following the guidelines in SOP 939 Revision 4.

The data contained in the following report have been reviewed and approved by the personnel listed below. In addition, ALS certifies that the analyses reported herein are true, complete and correct within the limits of the methods employed.



Megan Johnson
Inorganics Primary Data Reviewer

2/22/12
Date



Inorganics Final Data Reviewer

2/21/12
Date



Inorganic Data Reporting Qualifiers

The following qualifiers are used by the laboratory when reporting results of inorganic analyses.

- Concentration qualifier -- If the analyte was analyzed for but not detected a "U" is entered.
- QC qualifier -- Specified entries and their meanings are as follows:
 - N - Spiked sample recovery not within control limits.
 - * - Duplicate analysis (relative percent difference) not within control limits.
 - Z - Calibration spike recovery not within control limits.

ALS Environmental -- FC

Sample Number(s) Cross-Reference Table

OrderNum: 1202056

Client Name: NMED Hazardous Waste Bureau

Client Project Name: KAFB - BFF 1Q12

Client Project Number:

Client PO Number: 20-667-00-16004

Client Sample Number	Lab Sample Number	COC Number	Matrix	Date Collected	Time Collected
106070-A	1202056-1		WATER	06-Feb-12	11:01
106070-B	1202056-2		WATER	06-Feb-12	11:04
106070-C	1202056-3		WATER	06-Feb-12	11:06
106070-D	1202056-4		WATER	06-Feb-12	11:14
106070-E	1202056-5		WATER	06-Feb-12	11:17
106070-F	1202056-6		WATER	06-Feb-12	11:20
106070-G	1202056-7		WATER	06-Feb-12	11:23
106070-H	1202056-8		WATER	06-Feb-12	11:26
106070-I	1202056-9		WATER	06-Feb-12	11:29
106070-J	1202056-10		WATER	06-Feb-12	11:30



ALS Laboratory Group

225 Commerce Drive, Fort Collins, Colorado 80524
TF: (800) 443-1511 PH: (970) 490-1511 FX: (970) 490-1522

Chain-of-Custody

Form 2028

WORKORDER #

1202056

Lab ID	Field ID	Matrix	Sample Date	Sample Time	# Bottles	Pres.	QC	DATE		DISPOSAL	PAGE	
								TURNAROUND	DATE			
①	106070-A	W	2-6-12	11:01	3	HCl		8260	8260	8260	2-6-12	1 of 1
②	106070-B			11:04	3	HCl		8015	8015	8015	2-6-12	1 of 1
③	106070-C			11:06	3	HCl		8015	8015	8015	2-6-12	1 of 1
④	106070-D		11:14	11:14	1	-		8015	8015	8015	2-6-12	1 of 1
⑤	106070-E		11:20	11:20	1	H ₂ O ₂		8015	8015	8015	2-6-12	1 of 1
⑥	106070-F		11:20	11:20	1	-		8015	8015	8015	2-6-12	1 of 1
⑦	106070-G			11:23	1	H ₂ O ₂		8015	8015	8015	2-6-12	1 of 1
⑧	106070-H			11:26	1	Zn/NaOH		8015	8015	8015	2-6-12	1 of 1
⑨	106070-I			11:29	1	HNO ₃		8015	8015	8015	2-6-12	1 of 1
⑩	106070-J			11:30	1	HNO ₃		8015	8015	8015	2-6-12	1 of 1

*Time Zone: (Circle) EST CST MST PST Matrix: O = oil S = soil NS = non-soil solid W = water L = liquid E = extract F = filter

For metals or anions, please detail analytes below.

Comments:	RELINQUISHED BY	SIGNATURE	PRINTED NAME	DATE	TIME
	RECEIVED BY	Brian Salem	Brian Salem	2-6-12	15:00
	RELINQUISHED BY	Lawrence Schmitz	Lawrence Schmitz	2-11-12	09:45
	RECEIVED BY				
	RELINQUISHED BY				
	RECEIVED BY				



CONDITION OF SAMPLE UPON RECEIPT FORM

Client: NIMED
Project Manager: LRS

Workorder No: 1202056
Initials: LAS Date: 2/7/12

1. Does this project require any special handling in addition to standard Paragon procedures?		YES	<input checked="" type="radio"/> NO
2. Are custody seals on shipping containers intact?	NONE	<input checked="" type="radio"/> YES	NO
3. Are Custody seals on sample containers intact?	<input checked="" type="radio"/> NONE	YES	NO
4. Is there a COC (Chain-of-Custody) present or other representative documents?		<input checked="" type="radio"/> YES	NO
5. Are the COC and bottle labels complete and legible ?		<input checked="" type="radio"/> YES	NO
6. Is the COC in agreement with samples received? (IDs, dates, times, no. of samples, no. of containers, matrix, requested analyses, etc.)		<input checked="" type="radio"/> YES	NO
7. Were airbills / shipping documents present and/or removable?	DROP OFF	<input checked="" type="radio"/> YES	NO
8. Are all aqueous samples requiring preservation preserved correctly? (excluding volatiles)	N/A	<input checked="" type="radio"/> YES	NO
9. Are all aqueous non-preserved samples pH 4-9 ?	N/A	<input checked="" type="radio"/> YES	NO
10. Is there sufficient sample for the requested analyses?		<input checked="" type="radio"/> YES	NO
11. Were all samples placed in the proper containers for the requested analyses?		<input checked="" type="radio"/> YES	NO
12. Are all samples within holding times for the requested analyses?		<input checked="" type="radio"/> YES	NO
13. Were all sample containers received intact ? (not broken or leaking, etc.)		<input checked="" type="radio"/> YES	NO
14. Are all samples requiring no headspace (VOC, GRO, RSK/MEE, Rx CN/S, radon) headspace free? Size of bubble: ___ < green pea ___ > green pea	N/A	<input checked="" type="radio"/> YES	NO
15. Do perchlorate LCMS-MS samples have headspace ? (at least 1/3 of container required)	<input checked="" type="radio"/> N/A	YES	NO
16. Were samples checked for and free from the presence of residual chlorine ? (Applicable when PM has indicated samples are from a chlorinated water source; note if field preservation with sodium thiosulfate was not observed.)	<input checked="" type="radio"/> N/A	YES	NO
17. Were the samples shipped on ice ?		<input checked="" type="radio"/> YES	NO
18. Were cooler temperatures measured at 0.1-6.0°C? IR gun used*: <input checked="" type="radio"/> #2 <input type="radio"/> #4		RAD ONLY	<input checked="" type="radio"/> YES <input type="radio"/> NO
Cooler #: <u>1</u>			
Temperature (°C): <u>1.4</u>			
No. of custody seals on cooler: <u>2</u>			
External µR/hr reading: <u>15</u>			
Background µR/hr reading: <u>13</u>			
Were external µR/hr readings ≤ two times background and within DOT acceptance criteria? <input checked="" type="radio"/> YES <input type="radio"/> NO <input type="radio"/> NA (If no, see Form 008.)			

Additional Information: PROVIDE DETAILS BELOW FOR A NO RESPONSE TO ANY QUESTION ABOVE, EXCEPT #1 AND #16.

If applicable, was the client contacted? YES / NO / NA Contact: [Signature] Date/Time: _____

Project Manager Signature / Date: [Signature] 2/7/12

*IR Gun #2: Oakton, SN 29922500201-0066
Form 201r22.xls (6/1/09)

*IR Gun #4: Oakton, SN 2372220101-0002

1 From
 Date: 2/6/12
 Sender's FedEx Account Number: [blank]
 Sender's Name: Brian Salem
 Phone: 505 222-9576
 Company: NMEP / HWB
 Address: 5500 San Antonio NE
 City: Albuquerque State: NM ZIP: 87059

2 Your Internal Billing Reference
3 To
 Recipient's Name: Lanie Steere
 Phone: 970 490-1511
 Company: ALS Labs

Address: 225 Commerce Dr.
 We cannot deliver to P.O. boxes or P.O. ZIP codes.
 Address: [blank]
 City: Ft. Collins State: CO ZIP: 80624-2762

4 Express Package Service * To most locations.
 NOTE: Service order has changed. Please select carefully.
 Packages up to 150 lbs. For packages over 150 lbs., use the new FedEx Express Freight US Airbill.

Next Business Day
 06 FedEx First Overnight
 01 FedEx Priority Overnight
 05 FedEx Standard Overnight

2 or 3 Business Days
 49 NEW FedEx 2Day A.M.
 03 FedEx 2Day
 20 FedEx Express Saver


5 Packaging * Declared value limit \$500.
 06 FedEx Envelope* 02 FedEx Pak* 03 FedEx Box 04 FedEx Tube 01 Other

6 Special Handling and Delivery Signature Options
 03 SATURDAY DELIVERY
 No Signature Required
 10 Direct Signature
 34 Indirect Signature

Does this shipment contain dangerous goods?
 One box must be checked.
 No 04 Yes
 06 Dry Ice
 Cargo Aircraft Only

7 Payment Bill to:
 Enter FedEx Acct. No. or Credit Card No. below. Obtain recip. Acct. No.
 1 Sender Acct. No. in Section 1 will be billed. 2 Recipient 3 Third Party 4 Credit Card 5 Cash/Check

Total Packages: 1 Total Weight: 19.25 lbs. Credit Card Auth. [blacked out]
 *Our liability is limited to \$100 unless you declare a higher value. See the current FedEx Service Guide for details.



8768 4050 1517

BICARBONATE AS CaCO3

Method EPA310.1

Sample Results

Lab Name: ALS Environmental -- FC
Client Name: NMED Hazardous Waste Bureau
Client Project ID: KAFB - BFF 1Q12
Work Order Number: 1202056 **Final Volume:** 100 ml
Reporting Basis: As Received **Matrix:** WATER
Prep Method: NONE **Result Units:** MG/L
Analyst: Jason McNall

Client Sample ID	Lab ID	Date Collected	Date Prepared	Date Analyzed	Percent Moisture	Dilution Factor	Result	Reporting Limit	Flag	Sample Aliquot
106070-F	1202056-6	02/06/2012	02/15/2012	02/15/2012	N/A	1	160	10		50 ml

Comments:

1. ND or U = Not Detected at or above the client requested detection limit.

Data Package ID: ak1202056-1

CARBONATE AS CaCO3

Method EPA310.1

Sample Results

Lab Name: ALS Environmental -- FC
Client Name: NMED Hazardous Waste Bureau
Client Project ID: KAFB - BFF 1Q12
Work Order Number: 1202056 **Final Volume:** 100 ml
Reporting Basis: As Received **Matrix:** WATER
Prep Method: NONE **Result Units:** MG/L
Analyst: Jason McNall

Client Sample ID	Lab ID	Date Collected	Date Prepared	Date Analyzed	Percent Moisture	Dilution Factor	Result	Reporting Limit	Flag	Sample Aliquot
106070-F	1202056-6	02/06/2012	02/15/2012	02/15/2012	N/A	1	10	10	U	50 ml

Comments:

1. ND or U = Not Detected at or above the client requested detection limit.

Data Package ID: ak1202056-1

TOTAL ALKALINITY AS CaCO3

Method EPA310.1

Sample Results

Lab Name: ALS Environmental -- FC
Client Name: NMED Hazardous Waste Bureau
Client Project ID: KAFB - BFF 1Q12
Work Order Number: 1202056 **Final Volume:** 100 ml
Reporting Basis: As Received **Matrix:** WATER
Prep Method: NONE **Result Units:** MG/L
Analyst: Jason McNall

Client Sample ID	Lab ID	Date Collected	Date Prepared	Date Analyzed	Percent Moisture	Dilution Factor	Result	Reporting Limit	Flag	Sample Aliquot
106070-F	1202056-6	02/06/2012	02/15/2012	02/15/2012	N/A	1	160	10		50 ml

Comments:

1. ND or U = Not Detected at or above the client requested detection limit.

Data Package ID: ak1202056-1

AMMONIA AS N

Method EPA350.1

Sample Results

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Field ID:	106070-G
Lab ID:	1202056-7

Sample Matrix: WATER

% Moisture: N/A

Date Collected: 06-Feb-12

Date Extracted: 17-Feb-12

Date Analyzed: 17-Feb-12

Prep Method: NONE

Prep Batch: NH120217-1

QCBatchID: NH120217-1-1

Run ID: NH120217-1A

Cleanup: NONE

Basis: As Received

File Name: 0217NH.FDT

Analyst: Eric Allen Lin

Sample Aliquot: 5 ML

Final Volume: 5 ML

Result Units: MG/L

Clean DF: 1

Analysis ReqCode: 543

CASNO	Target Analyte	Dilution Factor	Result	Reporting Limit	Result Qualifier	EPA Qualifier
7664-41-7	AMMONIA AS N	1	0.1	0.1	U	

Data Package ID: *nh1202056-1*

Nitrate/Nitrite as N

Method EPA353.2 Revision 2.0

Sample Results

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Field ID:	106070-G
Lab ID:	1202056-7

Sample Matrix: WATER

% Moisture: N/A

Date Collected: 06-Feb-12

Date Extracted: 16-Feb-12

Date Analyzed: 16-Feb-12

Prep Method: NONE

Prep Batch: NN120216-1

QCBatchID: NN120216-1-1

Run ID: NN120216-1A

Cleanup: NONE

Basis: As Received

File Name: 0216NOX.FDT

Analyst: Eric Allen Lin

Sample Aliquot: 5 ML

Final Volume: 5 ML

Result Units: MG/L

Clean DF: 1

Analysis ReqCode: 542

CASNO	Target Analyte	Dilution Factor	Result	Reporting Limit	Result Qualifier	EPA Qualifier
1-005	NITRATE/NITRITE AS N	1	0.018	0.01		

Data Package ID: *nn1202056-1*

Sulfide

Method EPA376.1

Sample Results

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Field ID:	106070-H
Lab ID:	1202056-8

Sample Matrix: WATER

% Moisture: N/A

Date Collected: 06-Feb-12

Date Extracted: 08-Feb-12

Date Analyzed: 08-Feb-12

Prep Method: NONE

Prep Batch: S120208-1

QCBatchID: S120208-1-1

Run ID: s120208-1a

Cleanup: NONE

Basis: As Received

File Name:

Analyst: Jason McNal

Sample Aliquot: 200 ML

Final Volume: 200 ML

Result Units: MG/L

Clean DF: 1

Analysis ReqCode: 503

CASNO	Target Analyte	Dilution Factor	Result	Reporting Limit	Result Qualifier	EPA Qualifier
18496-25-8	SULFIDE	1	2	2	U	

Data Package ID: s1202056-1

Ion Chromatography

Method EPA300.0 Revision 2.1

Sample Results

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Field ID:	106070-F
Lab ID:	1202056-6

Sample Matrix: WATER

% Moisture: N/A

Date Collected: 06-Feb-12

Date Extracted: 08-Feb-12

Date Analyzed: 08-Feb-12

Prep Method: NONE

Prep Batch: IC120208-1

QCBatchID: IC120208-1-1

Run ID: IC120208-A1

Cleanup: NONE

Basis: As Received

File Name: 20208_027.DXD

Analyst: Eric Allen Lin

Sample Aliquot: 5 ML

Final Volume: 5 ML

Result Units: MG/L

Clean DF: 1

Analysis ReqCode: 524

CASNO	Target Analyte	Dilution Factor	Result	Reporting Limit	Result Qualifier	EPA Qualifier
16984-48-8	FLUORIDE AnalysisTime: 15:36	1	0.31	0.1		
16887-00-6	CHLORIDE AnalysisTime: 15:36	1	20	0.2		
24959-67-9	BROMIDE AnalysisTime: 15:36	1	0.29	0.2		
14808-79-8	SULFATE AnalysisTime: 15:36	1	46	1		

Data Package ID: ic1202056-1

BICARBONATE AS CaCO3

Method EPA310.1

Method Blank

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: AK120215-1MB

Sample Matrix: WATER

% Moisture: N/A

Prep Batch: AK120215-1

QCBatchID: AK120215-1-1

Run ID: ak120215-1a

Cleanup: NONE

Basis: N/A

Sample Aliquot: 100 ml

Final Volume: 100 ml

Result Units: MG/L

Lab ID	Date Prepared	Date Analyzed	Percent Moisture	Dilution Factor	Result	Reporting Limit	Flag
AK120215-1MB	2/15/2012	02/15/2012	N/A	1	5	5	U

Comments:

1. ND or U = Not Detected at or above the client requested detection limit.

Data Package ID: *ak1202056-1*

Date Printed: Monday, February 20, 2012

ALS Environmental -- FC

Page 1 of 3

LIMS Version: 6.562

CARBONATE AS CaCO3

Method EPA310.1

Method Blank

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: AK120215-1MB

Sample Matrix: WATER

% Moisture: N/A

Prep Batch: AK120215-1

QCBatchID: AK120215-1-1

Run ID: ak120215-1a

Cleanup: NONE

Basis: N/A

Sample Aliquot: 100 ml

Final Volume: 100 ml

Result Units: MG/L

Lab ID	Date Prepared	Date Analyzed	Percent Moisture	Dilution Factor	Result	Reporting Limit	Flag
AK120215-1MB	2/15/2012	02/15/2012	N/A	1	5	5	U

Comments:

1. ND or U = Not Detected at or above the client requested detection limit.

Data Package ID: *ak1202056-1*

Date Printed: Monday, February 20, 2012

ALS Environmental -- FC

Page 2 of 3

LIMS Version: 6.562

TOTAL ALKALINITY AS CaCO3

Method EPA310.1

Method Blank

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: AK120215-1MB

Sample Matrix: WATER

% Moisture: N/A

Prep Batch: AK120215-1

QCBatchID: AK120215-1-1

Run ID: ak120215-1a

Cleanup: NONE

Basis: N/A

Sample Aliquot: 100 ml

Final Volume: 100 ml

Result Units: MG/L

Lab ID	Date Prepared	Date Analyzed	Percent Moisture	Dilution Factor	Result	Reporting Limit	Flag
AK120215-1MB	2/15/2012	02/15/2012	N/A	1	5	5	U

Comments:

1. ND or U = Not Detected at or above the client requested detection limit.

Data Package ID: *ak1202056-1*

Date Printed: Monday, February 20, 2012

ALS Environmental -- FC

Page 3 of 3

LIMS Version: 6.562

TOTAL ALKALINITY AS CaCO3

Method EPA310.1

Laboratory Control Sample

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: AK120215-1LCS

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 02/15/2012

Date Analyzed: 02/15/2012

Prep Batch: AK120215-1

QCBatchID: AK120215-1-1

Run ID: ak120215-1a

Cleanup: NONE

Basis: N/A

Sample Aliquot: 100 ml

Final Volume: 100 ml

Result Units: MG/L

CASNO	Target Analyte	Spike Added	LCS Result	Reporting Limit	Result Qualifier	LCS % Rec.	Control Limits
	TOTAL ALKALINITY AS CaCO3	100	101	5		101	85 - 115

Data Package ID: ak1202056-1

Date Printed: Monday, February 20, 2012

ALS Environmental -- FC

LIMS Version: 6.562

Page 1 of 1

AMMONIA AS N

Method EPA350.1

Method Blank

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: NH120217-1MB

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 17-Feb-12

Date Analyzed: 17-Feb-12

Prep Method: NONE

Prep Batch: NH120217-1

QCBatchID: NH120217-1-1

Run ID: NH120217-1A

Cleanup: NONE

Basis: N/A

File Name: 0217NH.FDT

Sample Aliquot: 5 ml

Final Volume: 5 ml

Result Units: MG/L

Clean DF: 1

CASNO	Target Analyte	DF	Result	Reporting Limit	Result Qualifier	EPA Qualifier
7664-41-7	AMMONIA AS N	1	0.1	0.1	U	

Data Package ID: *nh1202056-1*

AMMONIA AS N

Method EPA350.1

Laboratory Control Sample

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: NH120217-1LCS

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 02/17/2012

Date Analyzed: 02/17/2012

Prep Method: NONE

Prep Batch: NH120217-1

QCBatchID: NH120217-1-1

Run ID: NH120217-1A

Cleanup: NONE

Basis: N/A

File Name: 0217NH.FDT

Sample Aliquot: 5 ml

Final Volume: 5 ml

Result Units: MG/L

Clean DF: 1

CASNO	Target Analyte	Spike Added	LCS Result	Reporting Limit	Result Qualifier	LCS % Rec.	Control Limits
7664-41-7	AMMONIA AS N	1	0.968	0.1		97	90 - 110%

Data Package ID: *nh1202056-1*

Date Printed: Monday, February 20, 2012

ALS Environmental -- FC

Page 1 of 1

LIMS Version: 6.562

Nitrate/Nitrite as N

Method EPA353.2 Revision 2.0

Method Blank

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: NN120216-1MB

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 16-Feb-12

Date Analyzed: 16-Feb-12

Prep Method: NONE

Prep Batch: NN120216-1

QCBatchID: NN120216-1-1

Run ID: NN120216-1A

Cleanup: NONE

Basis: N/A

File Name: 0216NOX.FDT

Sample Aliquot: 5 ml

Final Volume: 5 ml

Result Units: MG/L

Clean DF: 1

CASNO	Target Analyte	DF	Result	Reporting Limit	Result Qualifier	EPA Qualifier
1-005	NITRATE/NITRITE AS N	1	0.01	0.01	U	

Data Package ID: *nn1202056-1*

Nitrate/Nitrite as N

Method EPA353.2 Revision 2.0

Laboratory Control Sample

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: NN120216-1LCS	Sample Matrix: WATER % Moisture: N/A Date Collected: N/A Date Extracted: 02/16/2012 Date Analyzed: 02/16/2012 Prep Method: NONE	Prep Batch: NN120216-1 QCBatchID: NN120216-1-1 Run ID: NN120216-1A Cleanup: NONE Basis: N/A File Name: 0216NOX.FDT	Sample Aliquot: 5 ml Final Volume: 5 ml Result Units: MG/L Clean DF: 1
------------------------------	--	---	---

CASNO	Target Analyte	Spike Added	LCS Result	Reporting Limit	Result Qualifier	LCS % Rec.	Control Limits
1-005	NITRATE/NITRITE AS N	0.5	0.485	0.01		97	90 - 110%

Data Package ID: *nn1202056-1*

Sulfide

Method EPA376.1

Method Blank

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: S120208-1MB

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 08-Feb-12

Date Analyzed: 08-Feb-12

Prep Method: NONE

Prep Batch: S120208-1

QCBatchID: S120208-1-1

Run ID: s120208-1a

Cleanup: NONE

Basis: N/A

File Name:

Sample Aliquot: 200 ml

Final Volume: 200 ml

Result Units: MG/L

Clean DF: 1

CASNO	Target Analyte	DF	Result	Reporting Limit	Result Qualifier	EPA Qualifier
18496-25-8	SULFIDE	1	2	2	U	

Data Package ID: s1202056-1

Sulfide

Method EPA376.1

Laboratory Control Sample

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: S120208-1LCS

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 02/08/2012

Date Analyzed: 02/08/2012

Prep Method: NONE

Prep Batch: S120208-1

QCBatchID: S120208-1-1

Run ID: s120208-1a

Cleanup: NONE

Basis: N/A

File Name:

Sample Aliquot: 200 ml

Final Volume: 200 ml

Result Units: MG/L

Clean DF: 1

CASNO	Target Analyte	Spike Added	LCS Result	Reporting Limit	Result Qualifier	LCS % Rec.	Control Limits
18496-25-8	SULFIDE	11.5	11.8	2		103	80 - 120%

Data Package ID: s1202056-1

Date Printed: Monday, February 20, 2012

ALS Environmental -- FC

Page 1 of 1

LIMS Version: 6.562

Ion Chromatography

Method EPA300.0 Revision 2.1

Method Blank

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: IC120208-1MB

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 08-Feb-12

Date Analyzed: 08-Feb-12

Prep Method: NONE

Prep Batch: IC120208-1

QCBatchID: IC120208-1-1

Run ID: IC120208-A1

Cleanup: NONE

Basis: N/A

File Name: 20208_017.DXD

Sample Aliquot: 5 ml

Final Volume: 5 ml

Result Units: MG/L

Clean DF: 1

CASNO	Target Analyte	DF	Result	Reporting Limit	Result Qualifier	EPA Qualifier
16984-48-8	FLUORIDE	1	0.1	0.1	U	
16887-00-6	CHLORIDE	1	0.2	0.2	U	
24959-67-9	BROMIDE	1	0.2	0.2	U	
14808-79-8	SULFATE	1	1	1	U	

Data Package ID: ic1202056-1

Ion Chromatography

Method EPA300.0 Revision 2.1

Laboratory Control Sample

Lab Name: ALS Environmental -- FC

Work Order Number: 1202056

Client Name: NMED Hazardous Waste Bureau

ClientProject ID: KAFB - BFF 1Q12

Lab ID: IC120208-1LCS

Sample Matrix: WATER

% Moisture: N/A

Date Collected: N/A

Date Extracted: 02/08/2012

Date Analyzed: 02/08/2012

Prep Method: NONE

Prep Batch: IC120208-1

QCBatchID: IC120208-1-1

Run ID: IC120208-A1

Cleanup: NONE

Basis: N/A

File Name: 20208_018.DXD

Sample Aliquot: 5 ml

Final Volume: 5 ml

Result Units: MG/L

Clean DF: 1

CASNO	Target Analyte	Spike Added	LCS Result	Reporting Limit	Result Qualifier	LCS % Rec.	Control Limits
16984-48-8	FLUORIDE	2.5	2.48	0.1		99	90 - 110%
16887-00-6	CHLORIDE	5	5.08	0.2		102	90 - 110%
24959-67-9	BROMIDE	5	4.99	0.2		100	90 - 110%
14808-79-8	SULFATE	25	24.9	1		100	90 - 110%

Data Package ID: *ic1202056-1*

Date Printed: Monday, February 20, 2012

ALS Environmental -- FC

Page 1 of 1

LIMS Version: 6.562